Summary statistics - quantitative results

(Groups: manufacturer of kit)

Filter: Slovakia, minimal size of groups n = 5

EQA round: CC2/19 - Cystatin C

Dead line: 11.10.2019

RoM = robust average	AV = assigned value	Dmax = acceptable percent difference
SD = standard deviation	CRV = certified reference value	LL = lower limit
CV = coefficient of variation	RV = reference value	UL = upper limit
Ntot = total number of participants	CVE = consensus value from experts	Neva = number of evaluated participants
Nout = number of results excluded before calculation	CVP = consensus value from all participants	Nsuc = number of successful participants
	CVPG = consensus value from participants groups	Srel = success (relative)
	U_{AV} = expanded uncertainty of the assigned value (k = 2)	

Test [unit]	RoM	SD	CV [%]	N_{tot}	N_{out}	Comparability	AV	U_{AV}	D_{max}	LL	UL	N_{eva}	N_{suc}	S_{rel}
(106) Cystatin C [mg/L]				12								12	12	100%
Samples and groups														
Sample A	2,42	0,22	8,9	12		CVP	2,32	0,065	18%	1,9	2,74	12	12	100%
Other				12	0	2x 46, 2x 75, 2x 91, 2x 149, 2x 179, 2x 999						12		
Sample B	2,49	0,24	9,8	12		CVP	2,41	0,067	18%	1,97	2,85	12	12	100%
Other				12	0	2x 46, 2x 75, 2x 91, 2x 149, 2x 179, 2x 999						12		

st_kn_p

End of report

Printed: 16.10.2019