

RoM = robust average	AV = assigned value	Dmax = acceptable percent difference
SD = standard deviation	CRV = certified reference value	LL = lower limit
CV = coefficient of variation	RV = reference value	UL = upper limit
Ntot = total number of participants	CVE = consensus value from experts	Neva = number of evaluated participants
Nout = number of results excluded before calculation	CVP = consensus value from all participants	Nsuc = number of successful participants
	CVPG = consensus value from participants groups	Srel = success (relative)
	U _{AV} = expanded uncertainty of the assigned value (k = 2)	

Test	[unit]	Comparability												
		RoM	SD	CV [%]	N _{tot}	N _{out}	AV	U _{AV}	D _{max}	LL	UL	N _{eva}	N _{suc}	S _{rel}
Set 1														
(536) Prothrombin test (CoaguChek)					113							113	109	96%
Samples and groups														
Sample A1		1,20	0,00	0,00	113	CVP	1,2	20%	0,96	1,44	113	113	100%	
All results		1,20	0,00	0,00	113	0					113			
Sample B1		2,55	0,07	2,8	113	CVP	2,55	0,016	20%	2,04	3,06	113	109	96%
All results		2,55	0,07	2,8	113	0					113			
Set 3														
(543) Prothrombin test (microINR)					34							34	18	53%
Samples and groups														
Sample A3		1,32	0,33	25	34	CVP	1,32	0,14	20%	1,05	1,59	34	19	56%
All results		1,32	0,33	25	34	0					34			
Sample B3		1,95	0,24	12	34	CVP	1,95	0,10	20%	1,56	2,34	34	27	79%
All results		1,95	0,24	12	34	0					34			
Set 5														
(552) Prothrombin test (Xprecia)					19							19	19	100%
Samples and groups														
Sample A5		1,60	0,00	0,00	19	CVP	1,6	20%	1,28	1,92	19	19	100%	
All results		1,60	0,00	0,00	19	0					19			
Sample B5		2,75	0,062	2,2	19	CVP	2,75	0,035	20%	2,2	3,3	19	19	100%
All results		2,75	0,062	2,2	19	0					19			

st_kn_p

End of report

Printed: 07.10.2019