

RoM = robust average	AV = assigned value	Dmax = acceptable percent difference
SD = standard deviation	CRV = certified reference value	LL = lower limit
CV = coefficient of variation	RV = reference value	UL = upper limit
Ntot = total number of participants	CVE = consensus value from experts	Neva = number of evaluated participants
Nout = number of results excluded before calculation	CVP = consensus value from all participants	Nsuc = number of successful participants
	CVPG = consensus value from participants groups	Srel = success (relative)
	U <sub>AV</sub> = expanded uncertainty of the assigned value (k = 2)	

Test	[unit]						Comparability							
		RoM	SD	CV [%]	N <sub>tot</sub>	N <sub>out</sub>	AV	U <sub>AV</sub>	D <sub>max</sub>	LL	UL	N <sub>eva</sub>	N <sub>suc</sub>	S <sub>rel</sub>
<b>(150) WBC</b>					53							53	49	92%
Samples and groups	f.10(exp9)/L													
<b>Sample A</b>		6,01	0,28	4,6	53	CVP	6,03	0,038	15%	5,12	6,94	53	49	92%
All results		6,01	0,28	4,6	53	0						53		
<b>Sample B</b>		8,79	0,29	3,3	53	CVP	8,84	0,045	15%	7,51	10,2	53	53	100%
All results		8,79	0,29	3,3	53	0						53		
<b>(151) RBC</b>					53							53	53	100%
Samples and groups	f.10(exp12)/L													
<b>Sample A</b>		5,15	0,092	1,8	53	CVP	5,18	0,014	7%	4,81	5,55	53	53	100%
All results		5,15	0,092	1,8	53	0						53		
<b>Sample B</b>		4,49	0,078	1,7	53	CVP	4,51	0,011	7%	4,19	4,83	53	53	100%
All results		4,49	0,078	1,7	53	0						53		
<b>(152) Haemoglobin</b>					53							53	53	100%
Samples and groups	[g/L]													
<b>Sample A</b>		158	2,8	1,8	53	CVP	160	0,42	6%	150	170	53	53	100%
All results		158	2,8	1,8	53	0						53		
<b>Sample B</b>		134	2,1	1,6	53	CVP	135	0,31	6%	126	144	53	53	100%
All results		134	2,1	1,6	53	0						53		
<b>(153) HCT</b>					53							53	51	96%
Samples and groups	[-]													
<b>Sample A</b>		0,438	0,016	3,6	53							53	51	96%
(1) Automate		0,439	0,015	3,4	51	1	CVPG	0,448	,0027	10%	0,403	0,493	51	
Other					2	0							2	
							2x 2							
<b>Sample B</b>		0,371	0,014	3,7	53							53	51	96%
(1) Automate		0,372	0,013	3,4	51	1	CVPG	0,378	,0023	10%	0,34	0,416	51	
Other					2	0							2	
							2x 2							
<b>(154) MCV</b>					53							53	51	96%
Samples and groups	[fL]													
<b>Sample A</b>		84,8	2,8	3,4	53							53	51	96%
(1) Automate		85,0	2,7	3,2	51	0	CVPG	86,3	0,56	10%	77,6	95	51	
Other					2	0							2	
							2x 2							
<b>Sample B</b>		82,3	2,8	3,5	53							53	51	96%
(1) Automate		82,5	2,7	3,3	51	0	CVPG	83,7	0,54	10%	75,3	92,1	51	
Other					2	0							2	
							2x 2							
<b>(155) Platelets</b>					53							53	53	100%
Samples and groups	f.10(exp9)/L													
<b>Sample A</b>		75,7	5,7	7,5	53							53	53	100%
All results (without individual groups)		75,5	5,5	7,3	52	0	CVP	75,3	0,84	27%	54,9	95,7	52	
Other					1	0							1	
							1x 765							
<b>Sample B</b>		236	16	6,6	53							53	53	100%
All results		236	16	6,6	53	0	CVP	236	2,0	20%	188	284	53	
<b>(156) RDW</b>					43							41	39	95%
Samples and groups	[%]													
<b>Sample A</b>		11,9	2,3	20	43							41	40	98%
(749) Sysmex XE, XS, XT series		13,3	0,26	2,0	19	0	CVPG	13,2	0,13	10%	11,8	14,6	19	
(755) Sysmex KX series		8,66	0,52	6,0	10	0	CVPG	8,57	0,37	10%	7,71	9,43	10	
(772) Sysmex XN series		12,5	0,15	1,2	7	0	CVPG	12,6	0,051	10%	11,3	13,9	7	
Other					7	0							5	
							1x 737, 1x 765, 1x 768, 1x 771, 3x 773							
<b>Sample B</b>		12,4	2,2	18	43							41	39	95%
(749) Sysmex XE, XS, XT series		13,6	0,34	2,5	19	0	CVPG	13,5	0,14	10%	12,1	14,9	19	
(755) Sysmex KX series		9,23	0,72	7,8	10	0	CVPG	9,15	0,49	10%	8,23	10,1	10	
(772) Sysmex XN series		12,9	0,15	1,1	7	0	CVPG	13	0,059	10%	11,7	14,3	7	
Other					7	0							5	
							1x 737, 1x 765, 1x 768, 1x 771, 3x 773							
<b>(157) MPV</b>					47							47	47	100%
Samples and groups	[fL]													
<b>Sample A</b>		9,62	0,61	6,3	47							47	47	100%
(1) Automate; (63) Sysmex		9,67	0,50	5,2	40	0	CVPG	10	0,11	18%	8,2	11,8	40	
Other					7	0							7	
							3x 1/12, 2x 1/177, 1x 1/179, 1x 2/1							

Test	[unit]					Comparability									
		RoM	SD	CV [%]	N <sub>tot</sub>	N <sub>out</sub>	AV	U <sub>AV</sub>	D <sub>max</sub>	LL	UL	N <sub>eva</sub>	N <sub>suc</sub>	S <sub>rel</sub>	
<b>Sample B</b>		9,89	0,47	4,8	47							47	47	100%	
(1) Automate; (63) Sysmex		9,92	0,40	4,0	40	0	CVPG	10,2	0,077	18%	8,36	12,1	40		
Other					7	0							7		
														3x 1/12, 2x 1/177, 1x 1/179, 1x 2/1	
<b>(158) PDW [%]</b>					3							3	3	100%	
Samples and groups	[%]														
<b>Sample A</b>		30,6	18	59	3							3	3	100%	
Other					3	0						3			
														2x 12, 1x 179	
<b>Sample B</b>		25,8	12	47	3							3	3	100%	
Other					3	0						3			
														2x 12, 1x 179	
<b>(165) PDW [fL]</b>					38							38	35	92%	
Samples and groups	[fL]														
<b>Sample A</b>		11,3	0,75	6,6	38		CVP	11,2	0,19	15%	9,52	12,9	38	35	92%
All results		11,3	0,75	6,6	38	0							38		
<b>Sample B</b>		12,1	0,62	5,1	38		CVP	12	0,13	15%	10,2	13,8	38	38	100%
All results		12,1	0,62	5,1	38	0							38		
<b>(166) PDW [-]</b>					2							2	2	100%	
Samples and groups	[-]														
<b>Sample A</b>		19,2	2,3	12	2							2	2	100%	
Other					2	0						2			
														1x 1, 1x 177	
<b>Sample B</b>		17,3	0,80	4,6	2							2	2	100%	
Other					2	0						2			
														1x 1, 1x 177	
st_kn_p														End of report	
														Printed: 31.07.2019	