

Key: >>> ... marks correct (expected) result

> ... marks conditionally correct (acceptable) result

± ... marks the result not evaluated

N ... number of the results

Nrel ... relative number of the results

Test

Sample	Group	N	Result	N _{rel}
(261) Glucose				
A	All results	>>> 324	Negative, normal, <= 2,8 mmol/L	100 %
		1	11,1 - 28 mmol/L	0,31 %
----- Total per sample: n = 325 ----- Success: 100 % -----				
B	All results	2	Negative, normal, <= 2,8 mmol/L	0,62 %
		> 12	2,81 - 11 mmol/L	3,7 %
		>>> 144	11,1 - 28 mmol/L	44 %
		>>> 167	> 28 mmol/L	51 %
----- Total per sample: n = 325 ----- Success: 99 % -----				
Total success: 99 %				

(263) Ketone bodies

A	All results	>>> 324	Negative, <= 0,5 mmol/L	100 %
		1	3+, 4+	0,31 %
----- Total per sample: n = 325 ----- Success: 100 % -----				
B	All results	1	Negative, <= 0,5 mmol/L	0,31 %
		> 97	1+, 2+	30 %
		>>> 227	3+, 4+	70 %
----- Total per sample: n = 325 ----- Success: 100 % -----				
Total success: 100 %				

(264) Total protein

A	All results	>>> 323	Negative, < 0,3 g/L	100 %
		1	> 2,5 g/L	0,31 %
----- Total per sample: n = 324 ----- Success: 100 % -----				
B	All results	4	Negative, < 0,3 g/L	1,2 %
		8	0,3 - 0,5 g/L	2,5 %
		> 59	0,51 - 2,5 g/L	18 %
		>>> 253	> 2,5 g/L	78 %
----- Total per sample: n = 324 ----- Success: 96 % -----				
Total success: 96 %				

(266) Urobilinogen

A	All results	>>> 323	Negative, < 17 µmol/L	100 %
		1	1+, 2+	0,31 %
----- Total per sample: n = 324 ----- Success: 100 % -----				
B	All results	2	Negative, < 17 µmol/L	0,62 %
		> 67	1+, 2+	21 %
		>>> 255	3+, 4+	79 %
----- Total per sample: n = 324 ----- Success: 99 % -----				
Total success: 99 %				

(268) Bilirubin

A	All results	>>> 317	Negative	100 %
		1	3+, 4+	0,31 %
----- Total per sample: n = 318 ----- Success: 100 % -----				
B	All results	4	Negative	1,3 %
		> 29	1+, 2+	9,1 %
		>>> 285	3+, 4+	90 %
----- Total per sample: n = 318 ----- Success: 99 % -----				
Total success: 99 %				

(270) WBC

A	All results	>>> 278	Negative	99 %
		1	10 - 50 leu/µL	0,36 %
		1	51 - 250 leu/µL	0,36 %
----- Total per sample: n = 280 ----- Success: 99 % -----				
B	All results	2	Negative	0,71 %
		> 24	10 - 50 leu/µL	8,6 %
		>>> 111	51 - 250 leu/µL	40 %
		>>> 143	> 250 leu/µL	51 %
----- Total per sample: n = 280 ----- Success: 99 % -----				
Total success: 99 %				

(272) Blood

A	All results	1	Not specified	0,31 %
		>>> 318	Negative	98 %
		5	5 - 12 ery/µL	1,5 %
		1	> 50 ery/µL	0,31 %
----- Total per sample: n = 325 ----- Success: 98 % -----				
B	All results	1	Negative	0,31 %

Test

Sample	Group	N	Result	N _{rel}
B	All results	1	5 - 12 ery/ μ L	0,31 %
		> 25	13 - 50 ery/ μ L	7,7 %
		>>> 298	> 50 ery/ μ L	92 %
Total per sample: n = 325		Success: 99 %		
Total success: 98 %				

(273) pH

A	All results	>>> 275	5,0	85 %
		> 28	5,5	8,7 %
		> 16	6,0	5,0 %
		1	6,5	0,31 %
		3	7,0	0,93 %
Total per sample: n = 323		Success: 99 %		
Total success: 99 %				
B	All results	2	5,0	0,62 %
		> 13	6,0	4,0 %
		> 51	6,5	16 %
		>>> 173	7,0	54 %
		> 82	7,5	25 %
		> 2	8,0	0,62 %
Total per sample: n = 323		Success: 99 %		
Total success: 99 %				

(274) Nitrite

A	All results	>>> 284	Negative	100 %
		1	Positive	0,35 %
Total per sample: n = 285		Success: 100 %		
Total success: 99 %				
B	All results	2	Negative	0,70 %
		>>> 283	Positive	99 %
Total per sample: n = 285		Success: 99 %		
Total success: 99 %				

(275) hCG

A	All results	>>> 17	Negative	100 %
Total per sample: n = 17		Success: 100 %		
Total success: 94 %				
B	All results	1	Negative	5,9 %
		>>> 16	Positive	94 %
Total per sample: n = 17		Success: 94 %		
Total success: 94 %				