

RoM = robust average	AV = assigned value	Dmax = acceptable percent difference
SD = standard deviation	CRV = certified reference value	LL = lower limit
CV = coefficient of variation	RV = reference value	UL = upper limit
Ntot = total number of the participants	CVE = consensus value from experts	Neva = number of evaluated participants
Nout = number of results excluded before calculation	CVP = consensus value from all participants	Nsuc = number of successful participants
	CVPG = consensus value from participants groups	Srel = success (relative)
	U _{AV} = expanded uncertainty of the assigned value (k = 2)	

Test	[unit]	Test					Comparability							
		RoM	SD	CV [%]	N _{tot}	N _{out}	AV	U _{AV}	D _{max}	LL	UL	N _{eva}	N _{suc}	S _{rel}
(253) CD 3+ T lymphocytes					32							32	32	100%
Samples and groups	[%]													
Sample A		71,1	1,8	2,6	32	CVP	71,2	0,45	7%	66,2	76,2	32	32	100%
All results		71,1	1,8	2,6	32	0						32		
Sample B		65,1	1,4	2,1	32	CVP	65,1	0,35	7%	60,5	69,7	32	32	100%
All results		65,1	1,4	2,1	32	0						32		
(254) CD 4+ T lymphocytes					32							32	29	91%
Samples and groups	[%]													
Sample A		37,6	1,4	3,7	32	CVP	38	0,31	10%	34,2	41,8	32	32	100%
All results		37,6	1,4	3,7	32	0						32		
Sample B		24,8	1,2	4,7	32	CVP	25,2	0,26	10%	22,6	27,8	32	29	91%
All results		24,8	1,2	4,7	32	0						32		
(255) CD 8+ T lymphocytes					32							32	28	88%
Samples and groups	[%]													
Sample A		23,3	1,6	7,0	32	CVP	23,6	0,47	17%	19,5	27,7	32	28	88%
All results		23,3	1,6	7,0	32	0						32		
Sample B		27,8	1,3	4,8	32	CVP	28	0,34	17%	23,2	32,8	32	31	97%
All results		27,8	1,3	4,8	32	0						32		
(256) B lymphocytes					32							32	31	97%
Samples and groups	[%]													
Sample A		14,6	1,0	7,0	32	CVP	14,5	0,23	30%	10,1	18,9	32	31	97%
All results		14,6	1,0	7,0	32	0						32		
Sample B		19,1	1,3	6,7	32	CVP	19,4	0,27	30%	13,5	25,3	32	32	100%
All results		19,1	1,3	6,7	32	0						32		
(257) NK cells					32							32	25	78%
Samples and groups	[%]													
Sample A		12,4	2,2	18	32	CVP	12,3	0,46	26%	9,1	15,5	32	26	81%
All results		12,4	2,2	18	32	0						32		
Sample B		14,3	1,5	11	32	CVP	14	0,34	26%	10,3	17,7	32	30	94%
All results		14,3	1,5	11	32	0						32		