

EQA round: INRP2/21 - INR measurement on POCT

RoM = robust average	AV = assigned value	Dmax = acceptable percent difference
SD = standard deviation	CRV = certified reference value	LL = lower limit
CV = coefficient of variation	RV = reference value	UL = upper limit
Ntot = total number of the participants	CVE = consensus value from experts	Neva = number of evaluated participants
Nout = number of results excluded before calculation	CVP = consensus value from all participants	Nsuc = number of successful participants
	CVPG = consensus value from participants groups	Srel = success (relative)
	U _{AV} = expanded uncertainty of the assigned value (k = 2)	

Test	[unit]	Comparability												
		RoM	SD	CV [%]	N _{tot}	N _{out}	AV	U _{AV}	D _{max}	LL	UL	N _{eva}	N _{suc}	S _{rel}
Set 1														
(536) Prothrombin test (CoaguChek)					93							93	89	96%
Samples and groups														
Sample A1		2,48	0,066	2,6	93	CVP	2,48	0,017	20%	1,98	2,98	93	90	97%
All results		2,48	0,066	2,6	93	0						93		
Sample B1		3,18	0,092	2,9	93	CVP	3,18	0,023	20%	2,54	3,82	93	90	97%
All results		3,18	0,092	2,9	93	0						93		
st_kn_p														

End of report

Printed: 25.06.2021