

Key: >>> ... marks correct (expected) result
> ... marks conditionally correct (acceptable) result
± ... marks the result not evaluated

N ... number of the results
Nrel ... relative number of the results

Test

Sample	Group	N	Result	N _{rel}
(1001) DAT (direct Coombs test)				
1	All results	> 40	Negative	47 %
		>>> 42	Positive (1+)	49 %
		> 2	Positive (2+)	2,4 %
		> 1	Positive (3+)	1,2 %
Total per sample: n = 85		Success: 100 %		
2	All results	1	Negative	1,2 %
		> 2	Positive (2+)	2,4 %
		>>> 45	Positive (3+)	53 %
		> 37	Positive (4+)	44 %
Total per sample: n = 85		Success: 99 %		
Total success: 99 %				
(1002) Sensitization specificity anti-IgG				
1	All results	> 1	Not specified	2,4 %
		>>> 33	Negative	80 %
		> 6	Positive (1+)	15 %
		> 1	Positive (2+)	2,4 %
Total per sample: n = 41		Success: 100 %		
2	All results	> 1	Positive (2+)	2,4 %
		>>> 21	Positive (3+)	51 %
		> 19	Positive (4+)	46 %
Total per sample: n = 41		Success: 100 %		
Total success: 100 %				
(1003) Sensitization specificity anti-IgA				
1	All results	> 1	Not specified	3,2 %
		>>> 30	Negative	97 %
Total per sample: n = 31		Success: 100 %		
2	All results	>>> 31	Negative	100 %
Total per sample: n = 31		Success: 100 %		
Total success: 100 %				
(1004) Sensitization specificity anti-IgM				
1	All results	> 1	Not specified	3,2 %
		4	Negative	13 %
		> 18	Positive (1+)	58 %
		>>> 8	Positive (2+)	26 %
Total per sample: n = 31		Success: 87 %		
2	All results	>>> 31	Negative	100 %
Total per sample: n = 31		Success: 100 %		
Total success: 87 %				
(1005) Sensitization specificity anti-C3c				
1	All results	> 1	Not specified	3,1 %
		>>> 31	Negative	97 %
Total per sample: n = 32		Success: 100 %		
2	All results	>>> 32	Negative	100 %
Total per sample: n = 32		Success: 100 %		
Total success: 100 %				
(1006) Sensitization specificity anti-C3d				
1	All results	> 2	Not specified	5,4 %
		>>> 34	Negative	92 %
		> 1	Positive (1+)	2,7 %
Total per sample: n = 37		Success: 100 %		
2	All results	>>> 37	Negative	100 %
Total per sample: n = 37		Success: 100 %		
Total success: 100 %				
(1008) Col. agglut: Quantif. IgG-DAT				
1	All results	> 6	Not specified	25 %
		>>> 18	Negative	75 %
Total per sample: n = 24		Success: 100 %		
2	All results	> 1	1:10	4,2 %
		>>> 17	1:100	71 %
		> 6	1:300	25 %
Total per sample: n = 24		Success: 100 %		
Total success: 100 %				

EQA round: PAT1/21 - Direct Antiglobulin Test

Test		N	Result	N _{rel}
Sample	Group			
(1009) Col. agglut: Subclass IgG1 1:1				
1	All results	>	5 Not specified	23 %
		>>>	17 Negative	77 %
----- Total per sample: n = 22 -----		Success: 100 %		
2	All results		1 Negative	4,5 %
		>>>	17 Positive (1+)	77 %
		>	4 Positive (2+)	18 %
----- Total per sample: n = 22 -----		Success: 95 %		
Total success: 95 %				
(1010) Col. agglut: Subclass IgG1 1:100				
1	All results	>	5 Not specified	23 %
		>>>	17 Negative	77 %
----- Total per sample: n = 22 -----		Success: 100 %		
2	All results	>>>	22 Negative	100 %
----- Total per sample: n = 22 -----		Success: 100 %		
Total success: 100 %				
(1011) Col. agglut: Subclass IgG3 1:1				
1	All results	>	5 Not specified	23 %
		>>>	17 Negative	77 %
----- Total per sample: n = 22 -----		Success: 100 %		
2	All results	>>>	19 Negative	86 %
			2 Positive (1+)	9,1 %
			1 Positive (2+)	4,5 %
----- Total per sample: n = 22 -----		Success: 86 %		
Total success: 86 %				
(1012) Col. agglut: Subclass IgG3 1:100				
1	All results	>	5 Not specified	23 %
		>>>	17 Negative	77 %
----- Total per sample: n = 22 -----		Success: 100 %		
2	All results	>>>	22 Negative	100 %
----- Total per sample: n = 22 -----		Success: 100 %		
Total success: 100 %				
(1019) AB0 blood group				
1	All results	>>>	56 Blood group: B	100 %
----- Total per sample: n = 56 -----		Success: 100 %		
2	All results	>>>	56 Blood group: AB	100 %
----- Total per sample: n = 56 -----		Success: 100 %		
Total success: 100 %				
(1020) Rh D antigen				
1	All results		1 Not specified	1,8 %
		>>>	53 Negative (RhD-)	95 %
			2 w/v (RhD weak/variant)	3,6 %
----- Total per sample: n = 56 -----		Success: 95 %		
2	All results	>>>	56 Positive (RhD+)	100 %
----- Total per sample: n = 56 -----		Success: 100 %		
Total success: 95 %				
(1021) Antigen C				
1	All results	>>>	44 Negative	100 %
----- Total per sample: n = 44 -----		Success: 100 %		
2	All results	>>>	44 Positive	100 %
----- Total per sample: n = 44 -----		Success: 100 %		
Total success: 100 %				
(1022) Antigen c				
1	All results	>>>	44 Positive	100 %
----- Total per sample: n = 44 -----		Success: 100 %		
2	All results	>>>	44 Positive	100 %
----- Total per sample: n = 44 -----		Success: 100 %		
Total success: 100 %				
(1023) Antigen Cw				
1	All results	>>>	35 Negative	97 %
			1 Positive	2,8 %
----- Total per sample: n = 36 -----		Success: 97 %		
2	All results	>>>	35 Negative	97 %
			1 Positive	2,8 %
----- Total per sample: n = 36 -----		Success: 97 %		
Total success: 94 %				

EQA round: PAT1/21 - Direct Antiglobulin Test

Test		N	Result	N _{rel}
Sample	Group			
(1024) Antigen E				
1	All results	>>> 44	Negative	100 %
Total per sample: n = 44		Success: 100 %		
2	All results	>>> 44	Positive	100 %
Total per sample: n = 44		Success: 100 %		
Total success: 100 %				
(1025) Antigen e				
1	All results	>>> 44	Positive	100 %
Total per sample: n = 44		Success: 100 %		
2	All results	>>> 44	Positive	100 %
Total per sample: n = 44		Success: 100 %		
Total success: 100 %				
(1026) Antigen K				
1	All results	>>> 45	Negative	98 %
		1	Positive	2,2 %
Total per sample: n = 46		Success: 98 %		
2	All results	>>> 46	Negative	100 %
Total per sample: n = 46		Success: 100 %		
Total success: 98 %				
(1027) Antigen k				
1	All results	>>> 25	Positive	100 %
Total per sample: n = 25		Success: 100 %		
2	All results	>>> 24	Positive	96 %
		1	Not specified	4,0 %
Total per sample: n = 25		Success: 96 %		
Total success: 96 %				
(1028) Antigen M				
1	All results	>>> 25	Positive	100 %
Total per sample: n = 25		Success: 100 %		
2	All results	>>> 25	Positive	100 %
Total per sample: n = 25		Success: 100 %		
Total success: 100 %				
(1029) Antigen N				
1	All results	>>> 23	Negative	100 %
Total per sample: n = 23		Success: 100 %		
2	All results	>>> 23	Positive	100 %
Total per sample: n = 23		Success: 100 %		
Total success: 100 %				
(1030) Antigen S				
1	All results	>>> 25	Positive	100 %
Total per sample: n = 25		Success: 100 %		
2	All results	>>> 25	Positive	100 %
Total per sample: n = 25		Success: 100 %		
Total success: 100 %				
(1031) Antigen s				
1	All results	>>> 17	Positive	100 %
Total per sample: n = 17		Success: 100 %		
2	All results	>>> 17	Positive	100 %
Total per sample: n = 17		Success: 100 %		
Total success: 100 %				
(1032) Antigen P1				
1	All results	>>> 22	Positive	100 %
Total per sample: n = 22		Success: 100 %		
2	All results	>>> 20	Negative	91 %
		2	Positive	9,1 %
Total per sample: n = 22		Success: 91 %		
Total success: 91 %				
(1033) Antigen Jka				
1	All results	>>> 28	Positive	100 %
Total per sample: n = 28		Success: 100 %		
2	All results	>>> 27	Positive	96 %
		1	Not specified	3,6 %
Total per sample: n = 28		Success: 96 %		
Total success: 96 %				
(1034) Antigen Jkb				
1	All results	>>> 27	Positive	100 %

EQA round: PAT1/21 - Direct Antiglobulin Test

Test	Sample	Group	N	Result	N _{rel}
		Total per sample: n = 27	Success: 100 %		
2		All results	>>> 27	Positive	100 %
		Total per sample: n = 27	Success: 100 %		
Total success: 100 %					

(1035) Antigen Lea

1		All results	>>> 16	Negative	84 %
			3	Positive	16 %
		Total per sample: n = 19	Success: 84 %		
2		All results	>>> 16	Negative	84 %
			3	Positive	16 %
		Total per sample: n = 19	Success: 84 %		
Total success: 74 %					

(1036) Antigen Leb

1		All results	>>> 1	Negative	5,3 %
			18	Positive	95 %
		Total per sample: n = 19	Success: 95 %		
2		All results	>>> 3	Negative	16 %
			16	Positive	84 %
		Total per sample: n = 19	Success: 84 %		
Total success: 84 %					