

Key: >>> ... marks correct (expected) result
> ... marks conditionally correct (acceptable) result
± ... marks the result not evaluated

N ... number of the results
N_{rel} ... relative number of the results

Test

Sample	Group	N	Result	N _{rel}
(1001) DAT (direct Coombs test)				
1	All results	> 11	Positive (1+)	13 %
		>>> 55	Positive (2+)	65 %
		> 17	Positive (3+)	20 %
		> 2	Positive (4+)	2,4 %
Total per sample: n = 85		Success: 100 %		
2	All results	> 1	Positive (1+)	1,2 %
		> 6	Positive (2+)	7,1 %
		>>> 64	Positive (3+)	75 %
		> 14	Positive (4+)	16 %
Total per sample: n = 85		Success: 100 %		
Total success: 100 %				

(1002) Sensitization specificity anti-IgG

1	All results	1	Negative	2,5 %
		> 4	Positive (1+)	10 %
		>>> 27	Positive (2+)	68 %
		> 6	Positive (3+)	15 %
		> 2	Positive (4+)	5,0 %
Total per sample: n = 40		Success: 98 %		
2	All results	> 2	Positive (2+)	5,0 %
		>>> 32	Positive (3+)	80 %
		> 6	Positive (4+)	15 %
Total per sample: n = 40		Success: 100 %		
Total success: 98 %				

(1003) Sensitization specificity anti-IgA

1	All results	>>> 31	Negative	100 %
Total per sample: n = 31		Success: 100 %		
2	All results	>>> 31	Negative	100 %
Total per sample: n = 31		Success: 100 %		
Total success: 100 %				

(1004) Sensitization specificity anti-IgM

1	All results	>>> 31	Negative	100 %
Total per sample: n = 31		Success: 100 %		
2	All results	>>> 31	Negative	100 %
Total per sample: n = 31		Success: 100 %		
Total success: 100 %				

(1005) Sensitization specificity anti-C3c

1	All results	>>> 32	Negative	100 %
Total per sample: n = 32		Success: 100 %		
2	All results	>>> 32	Negative	100 %
Total per sample: n = 32		Success: 100 %		
Total success: 100 %				

(1006) Sensitization specificity anti-C3d

1	All results	>>> 36	Negative	100 %
Total per sample: n = 36		Success: 100 %		
2	All results	>>> 36	Negative	100 %
Total per sample: n = 36		Success: 100 %		
Total success: 100 %				

(1008) Col. agglut: Quantif. IgG-DAT

1	All results	> 14	1:10	61 %
		>>> 8	1:30	35 %
		> 1	1:300	4,3 %
Total per sample: n = 23		Success: 100 %		
2	All results	> 11	1:30	48 %
		>>> 11	1:100	48 %
		> 1	1:300	4,3 %
Total per sample: n = 23		Success: 100 %		
Total success: 100 %				

(1009) Col. agglut: Subclass IgG1 1:1

1	All results	>>> 22	Negative	100 %
Total per sample: n = 22		Success: 100 %		
2	All results	> 9	Negative	41 %
		>>> 10	Positive (1+)	45 %
		> 3	Positive (2+)	14 %

EQA round: PAT2/21 - Direct Antiglobulin Test

Test	Sample	Group	N	Result	N _{rel}
Total per sample: n = 22 ----- Success: 100 % -----					
Total success: 100 %					
(1010) Col. agglut: Subclass IgG1 1:100					
1	All results	>>>	21	Negative	100 %
----- Total per sample: n = 21 ----- Success: 100 % -----					
2	All results	>>>	21	Negative	100 %
----- Total per sample: n = 21 ----- Success: 100 % -----					
Total success: 100 %					
(1011) Col. agglut: Subclass IgG3 1:1					
1	All results	>>>	22	Negative	100 %
----- Total per sample: n = 22 ----- Success: 100 % -----					
2	All results	>>>	22	Negative	100 %
----- Total per sample: n = 22 ----- Success: 100 % -----					
Total success: 100 %					
(1012) Col. agglut: Subclass IgG3 1:100					
1	All results	>>>	21	Negative	100 %
----- Total per sample: n = 21 ----- Success: 100 % -----					
2	All results	>>>	21	Negative	100 %
----- Total per sample: n = 21 ----- Success: 100 % -----					
Total success: 100 %					
(1019) AB0 blood group					
1	All results	>>>	55	Blood group: 0	100 %
----- Total per sample: n = 55 ----- Success: 100 % -----					
2	All results	>>>	55	Blood group: A	100 %
----- Total per sample: n = 55 ----- Success: 100 % -----					
Total success: 100 %					
(1020) Rh D antigen					
1	All results	>>>	55	Negative (RhD-)	100 %
----- Total per sample: n = 55 ----- Success: 100 % -----					
2	All results	>>>	53	Negative (RhD-)	96 %
			1	Positive (RhD+)	1,8 %
			1	w/v (RhD weak/variant)	1,8 %
----- Total per sample: n = 55 ----- Success: 96 % -----					
Total success: 96 %					
(1021) Antigen C					
1	All results	>>>	44	Negative	100 %
----- Total per sample: n = 44 ----- Success: 100 % -----					
2	All results	>>>	44	Negative	100 %
----- Total per sample: n = 44 ----- Success: 100 % -----					
Total success: 100 %					
(1022) Antigen c					
1	All results	>>>	44	Positive	100 %
----- Total per sample: n = 44 ----- Success: 100 % -----					
2	All results	>>>	44	Positive	100 %
----- Total per sample: n = 44 ----- Success: 100 % -----					
Total success: 100 %					
(1023) Antigen Cw					
1	All results	>>>	36	Negative	100 %
----- Total per sample: n = 36 ----- Success: 100 % -----					
2	All results	>>>	36	Negative	100 %
----- Total per sample: n = 36 ----- Success: 100 % -----					
Total success: 100 %					
(1024) Antigen E					
1	All results	>>>	44	Negative	100 %
----- Total per sample: n = 44 ----- Success: 100 % -----					
2	All results	>>>	44	Negative	100 %
----- Total per sample: n = 44 ----- Success: 100 % -----					
Total success: 100 %					
(1025) Antigen e					
1	All results	>>>	44	Positive	100 %
----- Total per sample: n = 44 ----- Success: 100 % -----					
2	All results	>>>	44	Positive	100 %
----- Total per sample: n = 44 ----- Success: 100 % -----					
Total success: 100 %					
(1026) Antigen K					
1	All results	>>>	46	Negative	100 %

EQA round: PAT2/21 - Direct Antiglobulin Test

Test	Sample	Group	N	Result	N _{rel}
		Total per sample: n = 46	Success: 100 %		
2		All results	>>> 46	Positive	100 %
		Total per sample: n = 46	Success: 100 %		
Total success: 100 %					
(1027) Antigen k					
1		All results	>>> 23	Positive	100 %
		Total per sample: n = 23	Success: 100 %		
2		All results	>>> 23	Positive	100 %
		Total per sample: n = 23	Success: 100 %		
Total success: 100 %					
(1028) Antigen M					
1		All results	>>> 25	Positive	100 %
		Total per sample: n = 25	Success: 100 %		
2		All results	>>> 25	Positive	100 %
		Total per sample: n = 25	Success: 100 %		
Total success: 100 %					
(1029) Antigen N					
1		All results	>>> 23	Positive	100 %
		Total per sample: n = 23	Success: 100 %		
2		All results	>>> 23	Positive	100 %
		Total per sample: n = 23	Success: 100 %		
Total success: 100 %					
(1030) Antigen S					
1		All results	>>> 23	Negative	88 %
			3	Positive	12 %
		Total per sample: n = 26	Success: 88 %		
2		All results	>>> 26	Positive	100 %
		Total per sample: n = 26	Success: 100 %		
Total success: 88 %					
(1031) Antigen s					
1		All results	>>> 19	Positive	100 %
		Total per sample: n = 19	Success: 100 %		
2		All results	>>> 19	Positive	100 %
		Total per sample: n = 19	Success: 100 %		
Total success: 100 %					
(1032) Antigen P1					
1		All results	>>> 21	Negative	95 %
			1	Positive	4,5 %
		Total per sample: n = 22	Success: 95 %		
2		All results	>>> 22	Positive	100 %
		Total per sample: n = 22	Success: 100 %		
Total success: 95 %					
(1033) Antigen Jka					
1		All results	>>> 29	Positive	100 %
		Total per sample: n = 29	Success: 100 %		
2		All results	>>> 28	Positive	97 %
			1	Not specified	3,4 %
		Total per sample: n = 29	Success: 97 %		
Total success: 97 %					
(1034) Antigen Jkb					
1		All results	>>> 27	Positive	96 %
			1	Negative	3,6 %
		Total per sample: n = 28	Success: 96 %		
2		All results	>>> 28	Positive	100 %
		Total per sample: n = 28	Success: 100 %		
Total success: 96 %					
(1035) Antigen Lea					
1		All results	>>> 19	Negative	100 %
		Total per sample: n = 19	Success: 100 %		
2		All results	>>> 19	Positive	100 %
		Total per sample: n = 19	Success: 100 %		
Total success: 100 %					
(1036) Antigen Leb					
1		All results	>>> 19	Positive	100 %
		Total per sample: n = 19	Success: 100 %		
2		All results	>>> 18	Negative	95 %

EQA round: PAT2/21 - Direct Antiglobulin Test

Test	Sample	Group	N	Result	N _{rel}
	2	All results	1	Positive	5,3 %
----- Total per sample: n = 19 ----- Success: 95 % -----					
Total success: 95 %					

st_kl

End of report

Printed: 04.12.2021