

## EQA round: RET1/21 - Reticulocytes (Microscopy)

RoM = robust average	AV = assigned value	Dmax = acceptable percent difference
SD = standard deviation	CRV = certified reference value	LL = lower limit
CV = coefficient of variation	RV = reference value	UL = upper limit
Ntot = total number of the participants	CVE = consensus value from experts	Neva = number of evaluated participants
Nout = number of results excluded before calculation	CVP = consensus value from all participants	Nsuc = number of successful participants
	CVPG = consensus value from participants groups	Srel = success (relative)
	U <sub>AV</sub> = expanded uncertainty of the assigned value (k = 2)	

Test	[unit]	Comparability												
		RoM	SD	CV [%]	N <sub>tot</sub>	N <sub>out</sub>	AV	U <sub>AV</sub>	D <sub>max</sub>	LL	UL	N <sub>eva</sub>	N <sub>suc</sub>	S <sub>rel</sub>
<b>(211) Reticulocytes</b>					64							64	55	86%
Samples and groups	r.10(exp9)/L]													
<b>Sample A</b>		339	100	29	64	CVP	339	31	60%	135	543	64	61	95%
All results		339	100	29	64	0						64		
<b>Sample B</b>		69,2	22	32	64	CVP	69,2	6,7	60%	27,6	111	64	58	91%
All results		69,2	22	32	64	0						64		

st\_kn\_p

End of report

Printed: 22.06.2021