

## Summary statistics - quantitative results

(Groups: measurement principle)

Filter: Slovakia, minimal size of the groups n = 5

## EQA round: CRP2/22 - C-Reactive Protein

Deadline: 20.05.2022

RoM = robust average	AV = assigned value	Dmax = acceptable percent difference
SD = standard deviation	CRV = certified reference value	LL = lower limit
CV = coefficient of variation	RV = reference value	UL = upper limit
Ntot = total number of the participants	CVE = consensus value from experts	Neva = number of evaluated participants
Nout = number of results excluded before calculation	CVP = consensus value from all participants	Nsuc = number of successful participants
	CVPG = consensus value from participants groups	Srel = success (relative)
	U <sub>AV</sub> = expanded uncertainty of the assigned value (k = 2)	

Test	[unit]	Comparability												
		RoM	SD	CV [%]	N <sub>tot</sub>	N <sub>out</sub>	AV	U <sub>AV</sub>	D <sub>max</sub>	LL	UL	N <sub>eva</sub>	N <sub>suc</sub>	S <sub>rel</sub>
<b>(100) C-reactive protein</b>					27							27	27	100%
----- Samples and groups -----	[mg/L]													
<b>Sample A</b>		17,9	1,4	7,6	27	CVP	18,4	0,26	24%	13,9	22,9	27	27	100%
(2) Immunoturbidimetry		17,9	1,4	7,8	25	0						25		
Other					2	0						2		
						2x 3								
<b>Sample B</b>		66,6	4,8	7,2	27	CVP	63,4	0,76	24%	48,1	78,7	27	27	100%
(2) Immunoturbidimetry		66,5	4,5	6,8	25	0						25		
Other					2	0						2		
						2x 3								

st\_kn\_p

End of report

Printed: 25.05.2022