

Key:	>>> ... marks correct (expected) result	N ... number of the results
	> ... marks conditionally correct (acceptable) result	Nrel ... relative number of the results
	± ... marks the result not evaluated	

**Test**

Sample	Group	N	Result	N <sub>rel</sub>
<b>(268) Bilirubin</b>				
A	All results (without individual groups)	>>> 23	3+, 4+	82 %
	(223) Iris iChem Velocity	± 4	Negative	14 %
		± 1	1+, 2+	3,6 %
----- Total per sample: n = 28 ----- Success (only evaluated results): 100 % -----				
B	All results	>>> 28	Negative	100 %
----- Total per sample: n = 28 ----- Success: 100 % -----				
Total success (only evaluated tests): 100 %				

**(264) Total protein**

A	All results	2	0,3 - 0,5 g/L	7,1 %
		> 10	0,51 - 2,5 g/L	36 %
		>>> 16	> 2,5 g/L	57 %
----- Total per sample: n = 28 ----- Success: 93 % -----				
B	All results	2	Negative, < 0,3 g/L	7,1 %
		>>> 19	0,3 - 0,5 g/L	68 %
		>>> 6	0,51 - 2,5 g/L	21 %
		> 1	> 2,5 g/L	3,6 %
----- Total per sample: n = 28 ----- Success: 93 % -----				
Total success: 86 %				

**(261) Glucose**

A	All results	1	2,81 - 11 mmol/L	3,6 %
		> 6	11,1 - 28 mmol/L	21 %
		>>> 21	> 28 mmol/L	75 %
----- Total per sample: n = 28 ----- Success: 96 % -----				
B	All results	1	Negative, normal, <= 2,8 mmol/L	3,6 %
		> 2	2,81 - 11 mmol/L	7,1 %
		>>> 22	11,1 - 28 mmol/L	79 %
		> 3	> 28 mmol/L	11 %
----- Total per sample: n = 28 ----- Success: 96 % -----				
Total success: 96 %				

**(275) hCG**

A	All results	>>> 5	Positive	100 %
----- Total per sample: n = 5 ----- Success: 100 % -----				
B	All results	>>> 5	Negative	100 %
----- Total per sample: n = 5 ----- Success: 100 % -----				
Total success: 100 %				

**(263) Ketone bodies**

A	All results	> 18	1+, 2+	67 %
		>>> 9	3+, 4+	33 %
----- Total per sample: n = 27 ----- Success: 100 % -----				
B	All results	>>> 27	Negative, <= 0,5 mmol/L	100 %
----- Total per sample: n = 27 ----- Success: 100 % -----				
Total success: 100 %				

**(272) Blood**

A	All results	1	5 - 12 ery/μL	3,6 %
		> 7	13 - 50 ery/μL	25 %
		>>> 20	> 50 ery/μL	71 %
----- Total per sample: n = 28 ----- Success: 96 % -----				
B	All results	1	5 - 12 ery/μL	3,6 %
		> 1	13 - 50 ery/μL	3,6 %
		>>> 26	> 50 ery/μL	93 %
----- Total per sample: n = 28 ----- Success: 96 % -----				
Total success: 93 %				

**(270) WBC**

A	All results	> 10	51 - 250 leu/μL	38 %
		>>> 16	> 250 leu/μL	62 %
----- Total per sample: n = 26 ----- Success: 100 % -----				
B	All results	>>> 25	Negative	96 %
		1	> 250 leu/μL	3,8 %
----- Total per sample: n = 26 ----- Success: 96 % -----				
Total success: 96 %				

**(274) Nitrite**

A	All results	>>> 27	Positive	100 %
---	-------------	--------	----------	-------

## Test

Sample	Group	N	Result	N <sub>rel</sub>
Total per sample: n = 27		Success: 100 %		
B	All results	>>> 27	Negative	100 %
Total per sample: n = 27		Success: 100 %		
Total success: 100 %				

## (273) pH

A	All results	> 8	6,5	29 %
		>>> 13	7,0	46 %
		> 5	7,5	18 %
		> 2	8,0	7,1 %
Total per sample: n = 28		Success: 100 %		
B	All results	> 6	6,0	21 %
		>>> 20	6,5	71 %
		> 2	7,0	7,1 %
Total per sample: n = 28		Success: 100 %		
Total success: 100 %				

## (266) Urobilinogen

A	All results	> 12	1+, 2+	43 %
		>>> 16	3+, 4+	57 %
Total per sample: n = 28		Success: 100 %		
B	All results	>>> 28	Negative, < 17 µmol/L	100 %
Total per sample: n = 28		Success: 100 %		
Total success: 100 %				