Summary statistics - quantitative results

(Groups: manufacturer of kit)

Filter: Slovakia, minimal size of the groups n = 5

EQA round: FC1/22 - Calprotectin in Stool

Deadline: 04.03.2022

RoM = robust average	AV = assigned value	Dmax = acceptable percent difference
SD = standard deviation	CRV = certified reference value	LL = lower limit
CV = coefficient of variation	RV = reference value	UL = upper limit
Ntot = total number of the participants	CVE = consensus value from experts	Neva = number of evaluated participants
Nout = number of results excluded before calculation	CVP = consensus value from all participants	Nsuc = number of successful participants
	CVPG = consensus value from participants groups	Srel = success (relative)
	U_{AV} = expanded uncertainty of the assigned value (k = 2)	

Test [unit]	RoM	SD	CV [%]	N_{tot}	N_{out}	Comparability	AV	U_{AV}	D_{max}	LL	UL	N_{eva}	N_{suc}	S_{rel}
(223) Calprotectin [µg/g]				10								10	6	60%
Samples and groups														
Sample A	426	230	53	10		CVP	405	68	50%	202	608	10	6	60%
Other				10	0							10		
						2x 135, 1x 164, 1x 199, 1x 200, 1x 211, 4x 999								
Sample B	19,7	16	79	10		CVP	26,1	2,9	100%	0	52,2	10	9	90%
Other				10	1							10		
						2x 135, 1x 164, 1x 199, 1x 200, 1x 211, 4x 999								

st_kn_p

End of report

Printed: 15.03.2022