

RoM = robust average	AV = assigned value	Dmax = acceptable percent difference
SD = standard deviation	CRV = certified reference value	LL = lower limit
CV = coefficient of variation	RV = reference value	UL = upper limit
Ntot = total number of the participants	CVE = consensus value from experts	Neva = number of evaluated participants
Nout = number of results excluded before calculation	CVP = consensus value from all participants	Nsuc = number of successful participants
	CVPG = consensus value from participants groups	Srel = success (relative)
	U _{AV} = expanded uncertainty of the assigned value (k = 2)	

Test	[unit]	Test					Comparability							
		RoM	SD	CV [%]	N _{tot}	N _{out}	AV	U _{AV}	D _{max}	LL	UL	N _{eva}	N _{suc}	S _{rel}
(247) Gammopathies (quantity 1) (plasma)	[g/L]				14							14	10	71%
Samples and groups														
Sample A		19,8	2,6	13	14	CVP	20,6	0,79	33%	13,8	27,4	14	14	100%
All results		19,8	2,6	13	14	0						14		
Sample B		7,29	2,2	30	14	CVP	7,64	0,47	33%	5,11	10,2	14	10	71%
All results		7,29	2,2	30	14	0						14		
(279) Total protein (plasma)	[g/L]				13							13	13	100%
Samples and groups														
Sample A		68,8	1,7	2,5	13	CVP	69	0,37	8%	63,4	74,6	13	13	100%
All results		68,8	1,7	2,5	13	0						13		
Sample B		51,0	1,6	3,1	13	CVP	51,5	0,37	8%	47,3	55,7	13	13	100%
All results		51,0	1,6	3,1	13	0						13		
(314) FLC kappa (plasma)	[mg/L]				13							13	13	100%
Samples and groups														
Sample A		1,07	0,22	21	13	CVP	1,1	0,14	±3	0	4,1	13	13	100%
All results		1,07	0,22	21	13	0						13		
Sample B		4,83	0,56	12	13	CVP	4,7	0,40	±3	1,7	7,7	13	13	100%
All results		4,83	0,56	12	13	0						13		
(315) FLC lambda (plasma)	[mg/L]				13							13	11	85%
Samples and groups														
Sample A		144	27	19	13	CVP	149	6,5	50%	74,5	224	13	11	85%
All results		144	27	19	13	0						13		
Sample B		136	17	12	13	CVP	135	7,7	50%	67,5	203	13	11	85%
All results		136	17	12	13	0						13		

st_kn_p

End of report

Printed: 14.03.2022