

RoM = robust average	AV = assigned value	Dmax = acceptable percent difference
SD = standard deviation	CRV = certified reference value	LL = lower limit
CV = coefficient of variation	RV = reference value	UL = upper limit
Ntot = total number of the participants	CVE = consensus value from experts	Neva = number of evaluated participants
Nout = number of results excluded before calculation	CVP = consensus value from all participants	Nsuc = number of successful participants
	CVPG = consensus value from participants groups	Srel = success (relative)
	U _{AV} = expanded uncertainty of the assigned value (k = 2)	

Test	[unit]	Test					Comparability							
		RoM	SD	CV [%]	N _{tot}	N _{out}	AV	U _{AV}	D _{max}	LL	UL	N _{eva}	N _{suc}	S _{rel}
(247) Gammopathies (quantity 1) (plasma)	[g/L]				90							90	76	84%
Samples and groups														
Sample A		14,1	2,9	20	90	CVP	14,1	0,74	33%	9,44	18,8	90	76	84%
All results		14,1	2,9	20	90	0						90		
(279) Total protein (plasma)	[g/L]				85							85	83	98%
Samples and groups														
Sample A		62,5	1,8	2,9	85	CVP	62,5	0,48	8%	57,5	67,5	85	83	98%
All results		62,5	1,8	2,9	85	0						85		
(314) FLC kappa (plasma)	[mg/L]				58							58	51	88%
Samples and groups														
Sample A		154	32	21	58	CVP	154	10	50%	77	231	58	51	88%
All results		154	32	21	58	0						58		
(315) FLC lambda (plasma)	[mg/L]				58							58	55	95%
Samples and groups														
Sample A		4,24	0,82	19	58	CVP	4,24	0,26	±3	1,24	7,24	58	55	95%
All results		4,24	0,82	19	58	0						58		
(545) Gammopathies (quantity 1) (urine)	[g/L]				28							28	15	54%
Samples and groups														
Sample B		0,068	0,063	93	28	CVP	0,068	0,029	66%	0,023	0,113	28	15	54%
All results		0,068	0,063	93	28	0						28		
(546) Gammopathies (quantity 2) (urine)	[g/L]				10							10	8	80%
Samples and groups														
Sample B		0,212	0,073	34	10	CVP	0,212	0,057	66%	0,072	0,352	10	8	80%
All results		0,212	0,073	34	10	0						10		
(547) Gammopathies (quantity 3) (urine)	[g/L]				36							36	31	86%
Samples and groups														
Sample B		0,161	0,072	45	36	CVP	0,161	0,030	66%	0,054	0,268	36	31	86%
All results		0,161	0,072	45	36	0						36		
(550) Total protein (urine)	[g/L]				81							81	71	88%
Samples and groups														
Sample B		0,385	0,091	24	81	CVP	0,385	0,025	30%	0,269	0,501	81	71	88%
All results		0,385	0,091	24	81	0						81		