

RoM = robust average	AV = assigned value	Dmax = acceptable percent difference
SD = standard deviation	CRV = certified reference value	LL = lower limit
CV = coefficient of variation	RV = reference value	UL = upper limit
Ntot = total number of the participants	CVE = consensus value from experts	Neva = number of evaluated participants
Nout = number of results excluded before calculation	CVP = consensus value from all participants	Nsuc = number of successful participants
	CVPG = consensus value from participants groups	Srel = success (relative)
	U _{AV} = expanded uncertainty of the assigned value (k = 2)	

Test	[unit]	Comparability					Comparability							
		RoM	SD	CV [%]	N _{tot}	N _{out}	AV	U _{AV}	D _{max}	LL	UL	N _{eva}	N _{suc}	S _{rel}
(247) Gammopathies (quantity 1) (plasma)	[g/L]				15							15	13	87%
Samples and groups														
Sample A		12,6	2,2	17	15	CVP	14,1	0,74	33%	9,44	18,8	15	13	87%
All results		12,6	2,2	17	15	0						15		
(279) Total protein (plasma)	[g/L]				15							15	15	100%
Samples and groups														
Sample A		62,1	2,0	3,3	15	CVP	62,5	0,48	8%	57,5	67,5	15	15	100%
All results		62,1	2,0	3,3	15	0						15		
(314) FLC kappa (plasma)	[mg/L]				13							13	11	85%
Samples and groups														
Sample A		161	21	13	13	CVP	154	10	50%	77	231	13	11	85%
All results		161	21	13	13	0						13		
(315) FLC lambda (plasma)	[mg/L]				13							13	13	100%
Samples and groups														
Sample A		3,86	0,34	8,9	13	CVP	4,24	0,26	±3	1,24	7,24	13	13	100%
All results		3,86	0,34	8,9	13	0						13		
(545) Gammopathies (quantity 1) (urine)	[g/L]				6							6	6	100%
Samples and groups														
Sample B		0,065	0,053	82	6	CVP	0,068	0,029	66%	0,023	0,113	6	6	100%
All results		0,065	0,053	82	6	0						6		
(546) Gammopathies (quantity 2) (urine)	[g/L]				1							1	1	100%
Samples and groups														
Sample B		0,275			1	CVP	0,212	0,057	66%	0,072	0,352	1	1	100%
All results		0,275			1	0						1		
(547) Gammopathies (quantity 3) (urine)	[g/L]				10							10	7	70%
Samples and groups														
Sample B		0,158	0,12	76	10	CVP	0,161	0,030	66%	0,054	0,268	10	7	70%
All results		0,158	0,12	76	10	0						10		
(550) Total protein (urine)	[g/L]				15							15	13	87%
Samples and groups														
Sample B		0,413	0,12	28	15	CVP	0,385	0,025	30%	0,269	0,501	15	13	87%
All results		0,413	0,12	28	15	0						15		

st_kn_p

End of report

Printed: 07.10.2022