

Summary statistics - quantitative results

(Groups: manufacturer of kit)

Filter: minimal size of the groups n = 5

EQA round: LMWH2/22 - Low Molecular Weight Heparin

Deadline: 16.09.2022

RoM = robust average	AV = assigned value	Dmax = acceptable percent difference
SD = standard deviation	CRV = certified reference value	LL = lower limit
CV = coefficient of variation	RV = reference value	UL = upper limit
Ntot = total number of the participants	CVE = consensus value from experts	Neva = number of evaluated participants
Nout = number of results excluded before calculation	CVP = consensus value from all participants	Nsuc = number of successful participants
	CVPG = consensus value from participants groups	Srel = success (relative)
	U _{AV} = expanded uncertainty of the assigned value (k = 2)	

Test	[unit]	Comparability												
		RoM	SD	CV [%]	N _{tot}	N _{out}	AV	U _{AV}	D _{max}	LL	UL	N _{eva}	N _{suc}	S _{rel}
(103) Low Molecular Weight Heparin					135						135	131	97%	
Samples and groups														
Sample A		0,855	0,10	12	135	CVP	0,855	0,022	25%	0,641	1,07	135	133	99%
(38) IL		0,734	0,028	3,8	30	0						30		
(94) Stago		0,749	0,046	6,2	13	0						13		
(149) Siemens (Dade, BN, Dimension)		0,901	0,051	5,6	38	0						38		
(179) Siemens		0,908	0,053	5,8	28	0						28		
(193) HYPHEN BioMed		0,933	0,05	5,3	23	0						23		
Other					3	0						3		
-----							1x 0, 1x 12, 1x 102							
Sample B		1,47	0,16	11	135	CVP	1,47	0,034	25%	1,1	1,84	135	133	99%
(38) IL		1,27	0,041	3,2	30	0						30		
(94) Stago		1,41	0,10	7,4	13	0						13		
(149) Siemens (Dade, BN, Dimension)		1,52	0,031	2,0	38	0						38		
(179) Siemens		1,56	0,079	5,1	28	2						28		
(193) HYPHEN BioMed		1,60	0,089	5,5	23	0						23		
Other					3	0						3		
-----							1x 0, 1x 12, 1x 102							

st_kn_p

End of report

Printed: 28.09.2022