

## Summary statistics - quantitative results

(Groups: manufacturer of kit)

Filter: Slovakia, minimal size of the groups n = 5

## EQA round: LMWH2/22 - Low Molecular Weight Heparin

Deadline: 16.09.2022

RoM = robust average	AV = assigned value	Dmax = acceptable percent difference
SD = standard deviation	CRV = certified reference value	LL = lower limit
CV = coefficient of variation	RV = reference value	UL = upper limit
Ntot = total number of the participants	CVE = consensus value from experts	Neva = number of evaluated participants
Nout = number of results excluded before calculation	CVP = consensus value from all participants	Nsuc = number of successful participants
	CVPG = consensus value from participants groups	Srel = success (relative)
	U <sub>AV</sub> = expanded uncertainty of the assigned value (k = 2)	

Test	[unit]	Comparability												
		RoM	SD	CV [%]	N <sub>tot</sub>	N <sub>out</sub>	AV	U <sub>AV</sub>	D <sub>max</sub>	LL	UL	N <sub>eva</sub>	N <sub>suc</sub>	S <sub>rel</sub>
<b>(103) Low Molecular Weight Heparin</b>					17							17	15	88%
--- Samples and groups ---	[IU/mL]													
<b>Sample A</b>		0,884	0,08	9,0	17	CVP	0,855	0,022	25%	0,641	1,07	17	16	94%
Other					17	0						17		
							1x 0, 1x 12, 3x 38, 1x 94, 3x 149, 4x 179, 4x 193							
<b>Sample B</b>		1,49	0,18	12	17	CVP	1,47	0,034	25%	1,1	1,84	17	16	94%
Other					17	1						17		
							1x 0, 1x 12, 3x 38, 1x 94, 3x 149, 4x 179, 4x 193							

st\_kn\_p

End of report

Printed: 28.09.2022