

EQA round: CRP1/24 - C-Reactive Protein

Deadline: 9.2.2024

Setup: groups - M (measurement principle); Slovakia; minimal size of the groups n = 5

RoM = robust average

SD = standard deviation

CV = coefficient of variation

N_{tot} = total number of the resultsN_{out} = number of the results removed before calculation

AV = assigned value

CVP = consensus of all participants

U_{AV} = expanded uncertainty of the assigned value (k = 2)D_{max} = acceptable difference

LL = lower limit

UL = upper limit

N_{eva} = number of the results assessedN_{suc} = number of successful resultsS_{rel} = relative success

Test Sample Group	[unit]	RoM	SD	CV [%]	N _{tot}	N _{out}	Comparability					N _{eva}	N _{suc}	S _{rel} [%]
							AV	U _{AV}	D _{max}	LL	UL			
(100) C-reactive protein	[mg/L]				62							62	61	98
Sample A												62	61	98
(2) Immunoturbidimetry		17,2	1,4	8,1	62	0	CVP	17,4	0,22	24%	13,2	21,6	62	
Sample B												62	62	100
(2) Immunoturbidimetry		56,6	4,4	7,7	62	0	CVP	55,9	0,6	24%	42,4	69,4	62	