

Summary statistics - qualitative results**EQA round: DGP1/24 - Urine Strip Tests Analysis**

Deadline: 16.2.2024

Setup: groups - R (strip manufacturer)

AV = assigned value CVP = consensus of all participants >>> ... expected result > ... acceptable result ± ... result not assessed	N_{tot} = total number of the results N_{rel} = relative number of the results	N_{eva} = number of the results assessed N_{suc} = number of successful results S_{rel} = relative success
---	---	--

Test Sample Group	Frequency of the results				N_{eva}	N_{suc}	S_{rel} [%]
	AV	N_{tot}	N_{rel}	Result [%]			
(268) Bilirubin		291			291	289	99
Sample A		291			291	289	99
(5) Dialab	CVP >>>	3	1	Negative			
(12) Beckman Coulter	CVP >>>	9	3,1	Negative			
(46) Erba Lachema	CVP >>>	42	14	Negative			
		1	0,3	1+, 2+			
(49) BioVendor	CVP >>>	11	3,8	Negative			
(60) Roche	CVP >>>	25	8,6	Negative			
(63) Sysmex	CVP >>>	17	5,8	Negative			
		1	0,3	1+, 2+			
(109) Axis-Shield	CVP >>>	1	0,3	Negative			
(150) Arkray	CVP >>>	15	5,2	Negative			
(161) YD Diagnostics	CVP >>>	1	0,3	Negative			
(162) Siemens (Atellica)	CVP >>>	8	2,7	Negative			
(179) Siemens	CVP >>>	40	14	Negative			
(197) Analyticon	CVP >>>	2	0,7	Negative			
(201) Iris Diagnostics	CVP >>>	14	4,8	Negative			
(202) Dirui	CVP >>>	63	22	Negative			
(207) 77 Elektronika Kft.	CVP >>>	33	11	Negative			
(999) another manufacturer	CVP >>>	5	1,7	Negative			
Sample B		291			291	290	100
(5) Dialab	CVP >>>	3	1	Negative			
(12) Beckman Coulter	CVP >>>	9	3,1	Negative			
(46) Erba Lachema	CVP >>>	42	14	Negative			
		1	0,3	1+, 2+			
(49) BioVendor	CVP >>>	11	3,8	Negative			
(60) Roche	CVP >>>	25	8,6	Negative			
(63) Sysmex	CVP >>>	18	6,2	Negative			
(109) Axis-Shield	CVP >>>	1	0,3	Negative			
(150) Arkray	CVP >>>	15	5,2	Negative			
(161) YD Diagnostics	CVP >>>	1	0,3	Negative			
(162) Siemens (Atellica)	CVP >>>	8	2,7	Negative			
(179) Siemens	CVP >>>	40	14	Negative			
(197) Analyticon	CVP >>>	2	0,7	Negative			
(201) Iris Diagnostics	CVP >>>	14	4,8	Negative			
(202) Dirui	CVP >>>	63	22	Negative			
(207) 77 Elektronika Kft.	CVP >>>	33	11	Negative			
(999) another manufacturer	CVP >>>	5	1,7	Negative			
(264) Total protein		291			291	284	98
Sample A		291			291	289	99
(5) Dialab	CVP >>>	3	1	Negative, < 0,3 g/L			
(12) Beckman Coulter	CVP >>>	9	3,1	Negative, < 0,3 g/L			
(46) Erba Lachema	CVP >>>	44	15	Negative, < 0,3 g/L			
(49) BioVendor	CVP >>>	11	3,8	Negative, < 0,3 g/L			
(60) Roche	CVP >>>	25	8,6	Negative, < 0,3 g/L			
(63) Sysmex	CVP >>>	18	6,2	Negative, < 0,3 g/L			
(109) Axis-Shield	CVP >>>	1	0,3	Negative, < 0,3 g/L			
(150) Arkray	CVP >>>	14	4,8	Negative, < 0,3 g/L			
		1	0,3	0,3 - 0,5 g/L			
(161) YD Diagnostics	CVP >>>	1	0,3	Negative, < 0,3 g/L			
(162) Siemens (Atellica)	CVP >>>	8	2,7	Negative, < 0,3 g/L			
(179) Siemens	CVP >>>	40	14	Negative, < 0,3 g/L			
(197) Analyticon	CVP >>>	2	0,7	Negative, < 0,3 g/L			
(201) Iris Diagnostics	CVP >>>	13	4,5	Negative, < 0,3 g/L			
		1	0,3	0,3 - 0,5 g/L			
(202) Dirui	CVP >>>	63	22	Negative, < 0,3 g/L			
(207) 77 Elektronika Kft.	CVP >>>	33	11	Negative, < 0,3 g/L			
(999) another manufacturer	CVP >>>	4	1,4	Negative, < 0,3 g/L			
Sample B		291			291	286	98
(5) Dialab	CVP >>>	2	0,7	Negative, < 0,3 g/L			
		1	0,3	0,3 - 0,5 g/L			
(12) Beckman Coulter	CVP >>>	2	0,7	Negative, < 0,3 g/L			
		6	2,1	0,3 - 0,5 g/L			
		1	0,3	0,51 - 2,5 g/L			

Summary statistics - qualitative results**EQA round: DGP1/24 - Urine Strip Tests Analysis**

Deadline: 16.2.2024

Setup: groups - R (strip manufacturer)

Test Sample Group	AV	N _{tot}	N _{rel} [%]	Frequency of the results		N _{eva}	N _{suc}	S _{rel} [%]
				Result				
(46) Erba Lachema	CVP >>>	19	6,5	Negative, < 0,3 g/L				
	CVP >>>	24	8,2	0,3 - 0,5 g/L				
		1	0,3	0,51 - 2,5 g/L				
(49) BioVendor	CVP >>>	10	3,4	Negative, < 0,3 g/L				
	CVP >>>	1	0,3	0,3 - 0,5 g/L				
(60) Roche	CVP >>>	10	3,4	Negative, < 0,3 g/L				
	CVP >>>	15	5,2	0,3 - 0,5 g/L				
(63) Sysmex	CVP >>>	10	3,4	Negative, < 0,3 g/L				
	CVP >>>	8	2,7	0,3 - 0,5 g/L				
(109) Axis-Shield	CVP >>>	1	0,3	Negative, < 0,3 g/L				
(150) Arkray	CVP >>>	10	3,4	Negative, < 0,3 g/L				
	CVP >>>	5	1,7	0,3 - 0,5 g/L				
(161) YD Diagnostics	CVP >>>	1	0,3	0,3 - 0,5 g/L				
(162) Siemens (Atellica)	CVP >>>	4	1,4	Negative, < 0,3 g/L				
	CVP >>>	4	1,4	0,3 - 0,5 g/L				
(179) Siemens	CVP >>>	11	3,8	Negative, < 0,3 g/L				
	CVP >>>	28	9,6	0,3 - 0,5 g/L				
		1	0,3	0,51 - 2,5 g/L				
(197) Analyticon	CVP >>>	1	0,3	Negative, < 0,3 g/L				
	CVP >>>	1	0,3	0,3 - 0,5 g/L				
(201) Iris Diagnostics	CVP >>>	3	1	Negative, < 0,3 g/L				
	CVP >>>	11	3,8	0,3 - 0,5 g/L				
(202) Dirui	CVP >>>	23	7,9	Negative, < 0,3 g/L				
	CVP >>>	38	13	0,3 - 0,5 g/L				
		2	0,7	0,51 - 2,5 g/L				
(207) 77 Elektronika Kft.	CVP >>>	20	6,9	Negative, < 0,3 g/L				
	CVP >>>	13	4,5	0,3 - 0,5 g/L				
(999) another manufacturer	CVP >>>	2	0,7	Negative, < 0,3 g/L				
	CVP >>>	2	0,7	0,3 - 0,5 g/L				
(261) Glucose		292				292	265	91
Sample A		292				292	265	91
(5) Dialab	CVP >>>	1	0,3	2,81 - 11 mmol/L				
	CVP >>>	2	0,7	11,1 - 28 mmol/L				
(12) Beckman Coulter	CVP >>>	8	2,7	2,81 - 11 mmol/L				
	CVP >>>	1	0,3	11,1 - 28 mmol/L				
(46) Erba Lachema	CVP >>>	9	3,1	2,81 - 11 mmol/L				
	CVP >>>	22	7,5	11,1 - 28 mmol/L				
		13	4,5	> 28 mmol/L				
(49) BioVendor	CVP >>>	8	2,7	2,81 - 11 mmol/L				
	CVP >>>	2	0,7	11,1 - 28 mmol/L				
		1	0,3	> 28 mmol/L				
(60) Roche	CVP >>>	11	3,8	2,81 - 11 mmol/L				
	CVP >>>	14	4,8	11,1 - 28 mmol/L				
(63) Sysmex	CVP >>>	3	1	2,81 - 11 mmol/L				
	CVP >>>	15	5,1	11,1 - 28 mmol/L				
(109) Axis-Shield	CVP >>>	1	0,3	11,1 - 28 mmol/L				
(150) Arkray	CVP >>>	7	2,4	2,81 - 11 mmol/L				
	CVP >>>	7	2,4	11,1 - 28 mmol/L				
(161) YD Diagnostics	CVP >>>	1	0,3	11,1 - 28 mmol/L				
(162) Siemens (Atellica)		2	0,7	Negative, normal, <= 2,8 mmol/L				
	CVP >>>	3	1	2,81 - 11 mmol/L				
	CVP >>>	3	1	11,1 - 28 mmol/L				
(179) Siemens	CVP >>>	15	5,1	2,81 - 11 mmol/L				
	CVP >>>	25	8,6	11,1 - 28 mmol/L				
(197) Analyticon	CVP >>>	2	0,7	11,1 - 28 mmol/L				
(199) Eurolab Lambda	CVP >>>	1	0,3	11,1 - 28 mmol/L				
(201) Iris Diagnostics		2	0,7	Negative, normal, <= 2,8 mmol/L				
	CVP >>>	9	3,1	2,81 - 11 mmol/L				
	CVP >>>	2	0,7	11,1 - 28 mmol/L				
		1	0,3	> 28 mmol/L				
(202) Dirui		1	0,3	Negative, normal, <= 2,8 mmol/L				
	CVP >>>	25	8,6	2,81 - 11 mmol/L				
	CVP >>>	36	12	11,1 - 28 mmol/L				
		1	0,3	> 28 mmol/L				
(207) 77 Elektronika Kft.	CVP >>>	18	6,2	2,81 - 11 mmol/L				
	CVP >>>	10	3,4	11,1 - 28 mmol/L				
		5	1,7	> 28 mmol/L				
(999) another manufacturer	CVP >>>	1	0,3	2,81 - 11 mmol/L				
	CVP >>>	3	1	11,1 - 28 mmol/L				

Summary statistics - qualitative results**EQA round: DGP1/24 - Urine Strip Tests Analysis**

Deadline: 16.2.2024

Setup: groups - R (strip manufacturer)

Test Sample Group	AV	N _{tot}	N _{rel} [%]	Frequency of the results		Success N _{eva} N _{suc} S _{rel} [%]
				Result		
		1	0,3	> 28 mmol/L		
Sample B		292				292 289 99
(5) Dialab	CVP >>>	3	1	Negative, normal, <= 2,8 mmol/L		
(12) Beckman Coulter	CVP >>>	9	3,1	Negative, normal, <= 2,8 mmol/L		
(46) Erba Lachema	CVP >>>	44	15	Negative, normal, <= 2,8 mmol/L		
(49) BioVendor	CVP >>>	11	3,8	Negative, normal, <= 2,8 mmol/L		
(60) Roche	CVP >>>	25	8,6	Negative, normal, <= 2,8 mmol/L		
(63) Sysmex	CVP >>>	18	6,2	Negative, normal, <= 2,8 mmol/L		
(109) Axis-Shield	CVP >>>	1	0,3	Negative, normal, <= 2,8 mmol/L		
(150) Arkay	CVP >>>	14	4,8	Negative, normal, <= 2,8 mmol/L		
(161) YD Diagnostics	CVP >>>	1	0,3	Negative, normal, <= 2,8 mmol/L		
(162) Siemens (Atellica)	CVP >>>	7	2,4	Negative, normal, <= 2,8 mmol/L		
		1	0,3	2,81 - 11 mmol/L		
(179) Siemens	CVP >>>	40	14	Negative, normal, <= 2,8 mmol/L		
(197) Analyticon	CVP >>>	2	0,7	Negative, normal, <= 2,8 mmol/L		
(199) Eurolab Lambda	CVP >>>	1	0,3	Negative, normal, <= 2,8 mmol/L		
(201) Iris Diagnostics	CVP >>>	13	4,5	Negative, normal, <= 2,8 mmol/L		
		1	0,3	2,81 - 11 mmol/L		
(202) Dirui	CVP >>>	63	22	Negative, normal, <= 2,8 mmol/L		
(207) 77 Elektronika Kft.	CVP >>>	33	11	Negative, normal, <= 2,8 mmol/L		
(999) another manufacturer	CVP >>>	4	1,4	Negative, normal, <= 2,8 mmol/L		
		1	0,3	2,81 - 11 mmol/L		
(275) hCG		11				11 11 100
Sample A		11				11 11 100
(5) Dialab	CVP >>>	2	18	Negative		
(17) Biogema	CVP >>>	3	27	Negative		
(115) BioSystems	CVP >>>	1	9,1	Negative		
(179) Siemens	CVP >>>	1	9,1	Negative		
(202) Dirui	CVP >>>	1	9,1	Negative		
(999) another manufacturer	CVP >>>	3	27	Negative		
Sample B		11				11 11 100
(5) Dialab	CVP >>>	2	18	Positive		
(17) Biogema	CVP >>>	3	27	Positive		
(115) BioSystems	CVP >>>	1	9,1	Positive		
(179) Siemens	CVP >>>	1	9,1	Positive		
(202) Dirui	CVP >>>	1	9,1	Positive		
(999) another manufacturer	CVP >>>	3	27	Positive		
(263) Ketone bodies		291				291 290 100
Sample A		291				291 290 100
(5) Dialab	CVP >>>	3	1	Negative, <= 0,5 mmol/L		
(12) Beckman Coulter	CVP >>>	9	3,1	Negative, <= 0,5 mmol/L		
(46) Erba Lachema	CVP >>>	43	15	Negative, <= 0,5 mmol/L		
		1	0,3	1+, 2+		
(49) BioVendor	CVP >>>	11	3,8	Negative, <= 0,5 mmol/L		
(60) Roche	CVP >>>	25	8,6	Negative, <= 0,5 mmol/L		
(63) Sysmex	CVP >>>	18	6,2	Negative, <= 0,5 mmol/L		
(109) Axis-Shield	CVP >>>	1	0,3	Negative, <= 0,5 mmol/L		
(150) Arkay	CVP >>>	15	5,2	Negative, <= 0,5 mmol/L		
(161) YD Diagnostics	CVP >>>	1	0,3	Negative, <= 0,5 mmol/L		
(162) Siemens (Atellica)	CVP >>>	8	2,7	Negative, <= 0,5 mmol/L		
(179) Siemens	CVP >>>	39	13	Negative, <= 0,5 mmol/L		
(197) Analyticon	CVP >>>	2	0,7	Negative, <= 0,5 mmol/L		
(201) Iris Diagnostics	CVP >>>	14	4,8	Negative, <= 0,5 mmol/L		
(202) Dirui	CVP >>>	63	22	Negative, <= 0,5 mmol/L		
(207) 77 Elektronika Kft.	CVP >>>	33	11	Negative, <= 0,5 mmol/L		
(999) another manufacturer	CVP >>>	5	1,7	Negative, <= 0,5 mmol/L		
Sample B		291				291 290 100
(5) Dialab	CVP >>>	3	1	Negative, <= 0,5 mmol/L		
(12) Beckman Coulter	CVP >>>	9	3,1	Negative, <= 0,5 mmol/L		
(46) Erba Lachema	CVP >>>	43	15	Negative, <= 0,5 mmol/L		
		1	0,3	1+, 2+		
(49) BioVendor	CVP >>>	11	3,8	Negative, <= 0,5 mmol/L		
(60) Roche	CVP >>>	25	8,6	Negative, <= 0,5 mmol/L		
(63) Sysmex	CVP >>>	18	6,2	Negative, <= 0,5 mmol/L		
(109) Axis-Shield	CVP >>>	1	0,3	Negative, <= 0,5 mmol/L		
(150) Arkay	CVP >>>	15	5,2	Negative, <= 0,5 mmol/L		
(161) YD Diagnostics	CVP >>>	1	0,3	Negative, <= 0,5 mmol/L		
(162) Siemens (Atellica)	CVP >>>	8	2,7	Negative, <= 0,5 mmol/L		
(179) Siemens	CVP >>>	39	13	Negative, <= 0,5 mmol/L		

Summary statistics - qualitative results**EQA round: DGP1/24 - Urine Strip Tests Analysis**

Deadline: 16.2.2024

Setup: groups - R (strip manufacturer)

Test Sample Group	AV	N _{tot}	N _{rel} [%]	Frequency of the results		N _{eva}	N _{suc}	S _{rel} [%]
				Result				
(197) Analyticon	CVP >>>	2	0,7	Negative, <= 0,5 mmol/L				
(201) Iris Diagnostics	CVP >>>	14	4,8	Negative, <= 0,5 mmol/L				
(202) Dirui	CVP >>>	63	22	Negative, <= 0,5 mmol/L				
(207) 77 Elektronika Kft.	CVP >>>	33	11	Negative, <= 0,5 mmol/L				
(999) another manufacturer	CVP >>>	5	1,7	Negative, <= 0,5 mmol/L				
(272) Blood		292				292	280	96
Sample A		292				292	285	98
(5) Dialab	>	1	0,3	13 - 50 ery/µL				
	CVP >>>	2	0,7	> 50 ery/µL				
(12) Beckman Coulter	>	1	0,3	13 - 50 ery/µL				
	CVP >>>	8	2,7	> 50 ery/µL				
(46) Erba Lachema		3	1	5 - 12 ery/µL				
	>	14	4,8	13 - 50 ery/µL				
	CVP >>>	27	9,2	> 50 ery/µL				
(49) BioVendor	CVP >>>	11	3,8	> 50 ery/µL				
(60) Roche		1	0,3	5 - 12 ery/µL				
	>	1	0,3	13 - 50 ery/µL				
	CVP >>>	23	7,9	> 50 ery/µL				
(63) Sysmex	>	1	0,3	13 - 50 ery/µL				
	CVP >>>	17	5,8	> 50 ery/µL				
(109) Axis-Shield	CVP >>>	1	0,3	> 50 ery/µL				
(150) Arkray		1	0,3	5 - 12 ery/µL				
	>	2	0,7	13 - 50 ery/µL				
	CVP >>>	12	4,1	> 50 ery/µL				
(161) YD Diagnostics	CVP >>>	1	0,3	> 50 ery/µL				
(162) Siemens (Atellica)	>	2	0,7	13 - 50 ery/µL				
	CVP >>>	6	2,1	> 50 ery/µL				
(179) Siemens	>	5	1,7	13 - 50 ery/µL				
	CVP >>>	35	12	> 50 ery/µL				
(197) Analyticon	CVP >>>	2	0,7	> 50 ery/µL				
(201) Iris Diagnostics		2	0,7	5 - 12 ery/µL				
	>	2	0,7	13 - 50 ery/µL				
	CVP >>>	10	3,4	> 50 ery/µL				
(202) Dirui	>	6	2,1	13 - 50 ery/µL				
	CVP >>>	57	20	> 50 ery/µL				
(207) 77 Elektronika Kft.	>	1	0,3	13 - 50 ery/µL				
	CVP >>>	32	11	> 50 ery/µL				
(999) another manufacturer	>	1	0,3	13 - 50 ery/µL				
	CVP >>>	4	1,4	> 50 ery/µL				
Sample B		292				292	286	98
(5) Dialab	CVP >>>	2	0,7	Negative				
	>	1	0,3	5 - 12 ery/µL				
(12) Beckman Coulter	CVP >>>	4	1,4	Negative				
	>	5	1,7	5 - 12 ery/µL				
(46) Erba Lachema	CVP >>>	30	10	Negative				
	>	14	4,8	5 - 12 ery/µL				
(49) BioVendor	CVP >>>	5	1,7	Negative				
	>	6	2,1	5 - 12 ery/µL				
(60) Roche	CVP >>>	22	7,5	Negative				
	>	3	1	5 - 12 ery/µL				
(63) Sysmex	CVP >>>	18	6,2	Negative				
(109) Axis-Shield	>	1	0,3	5 - 12 ery/µL				
(150) Arkray	CVP >>>	13	4,5	Negative				
	>	2	0,7	5 - 12 ery/µL				
(161) YD Diagnostics	>	1	0,3	5 - 12 ery/µL				
(162) Siemens (Atellica)	CVP >>>	8	2,7	Negative				
(179) Siemens	CVP >>>	30	10	Negative				
	>	8	2,7	5 - 12 ery/µL				
		2	0,7	13 - 50 ery/µL				
(197) Analyticon	CVP >>>	2	0,7	Negative				
(201) Iris Diagnostics	CVP >>>	4	1,4	Negative				
	>	8	2,7	5 - 12 ery/µL				
		1	0,3	13 - 50 ery/µL				
		1	0,3	> 50 ery/µL				
(202) Dirui	CVP >>>	47	16	Negative				
	>	15	5,1	5 - 12 ery/µL				
		1	0,3	> 50 ery/µL				
(207) 77 Elektronika Kft.	CVP >>>	8	2,7	Negative				
	>	24	8,2	5 - 12 ery/µL				

Summary statistics - qualitative results**EQA round: DGP1/24 - Urine Strip Tests Analysis**

Deadline: 16.2.2024

Setup: groups - R (strip manufacturer)

Test Sample Group	AV	N _{tot}	N _{rel} [%]	Frequency of the results		N _{eva}	N _{suc}	S _{rel} [%]
				Result				
(999) another manufacturer	CVP >>>	5	0,3 1,7	13 - 50 ery/µL Negative				
(270) WBC		262				262	260	99
Sample A		262				262	261	100
(5) Dialab	CVP >>>	2	0,8	Negative				
(12) Beckman Coulter		1	0,4	Not specified				
(46) Erba Lachema	CVP >>>	8	3,1	Negative				
(49) BioVendor	CVP >>>	11	4,2	Negative				
(60) Roche	CVP >>>	24	9,2	Negative				
(63) Sysmex	CVP >>>	18	6,9	Negative				
(109) Axis-Shield	CVP >>>	1	0,4	Negative				
(150) Arkay	CVP >>>	15	5,7	Negative				
(162) Siemens (Atellica)	CVP >>>	8	3,1	Negative				
(179) Siemens	CVP >>>	39	15	Negative				
(197) Analyticon	CVP >>>	2	0,8	Negative				
(201) Iris Diagnostics	CVP >>>	14	5,3	Negative				
(202) Dirui	CVP >>>	63	24	Negative				
(207) 77 Elektronika Kft.	CVP >>>	33	13	Negative				
(999) another manufacturer	CVP >>>	5	1,9	Negative				
Sample B		262				262	261	100
(5) Dialab	CVP >>>	2	0,8	Negative				
(12) Beckman Coulter	CVP >>>	9	3,4	Negative				
(46) Erba Lachema	CVP >>>	18	6,9	Negative				
(49) BioVendor	CVP >>>	11	4,2	Negative				
(60) Roche	CVP >>>	24	9,2	Negative				
(63) Sysmex	CVP >>>	18	6,9	Negative				
(109) Axis-Shield	CVP >>>	1	0,4	Negative				
(150) Arkay	CVP >>>	15	5,7	Negative				
(162) Siemens (Atellica)	CVP >>>	8	3,1	Negative				
(179) Siemens	CVP >>>	39	15	Negative				
(197) Analyticon	CVP >>>	2	0,8	Negative				
(201) Iris Diagnostics	CVP >>>	14	5,3	Negative				
(202) Dirui	CVP >>>	63	24	Negative				
(207) 77 Elektronika Kft.	CVP >>>	33	13	Negative				
(999) another manufacturer	CVP >>>	4	1,5	Negative				
		1	0,4	10 - 50 leu/µL				
(274) Nitrite		263				263	262	100
Sample A		263				263	263	100
(5) Dialab	CVP >>>	3	1,1	Negative				
(9) Bayer	CVP >>>	1	0,4	Negative				
(12) Beckman Coulter	CVP >>>	9	3,4	Negative				
(46) Erba Lachema	CVP >>>	19	7,2	Negative				
(49) BioVendor	CVP >>>	11	4,2	Negative				
(60) Roche	CVP >>>	23	8,7	Negative				
(63) Sysmex	CVP >>>	18	6,8	Negative				
(109) Axis-Shield	CVP >>>	1	0,4	Negative				
(150) Arkay	CVP >>>	15	5,7	Negative				
(162) Siemens (Atellica)	CVP >>>	8	3	Negative				
(179) Siemens	CVP >>>	39	15	Negative				
(197) Analyticon	CVP >>>	2	0,8	Negative				
(201) Iris Diagnostics	CVP >>>	14	5,3	Negative				
(202) Dirui	CVP >>>	63	24	Negative				
(207) 77 Elektronika Kft.	CVP >>>	32	12	Negative				
(999) another manufacturer	CVP >>>	5	1,9	Negative				
Sample B		263				263	262	100
(5) Dialab	CVP >>>	3	1,1	Negative				
(9) Bayer	CVP >>>	1	0,4	Negative				
(12) Beckman Coulter	CVP >>>	9	3,4	Negative				
(46) Erba Lachema	CVP >>>	19	7,2	Negative				
(49) BioVendor	CVP >>>	11	4,2	Negative				
(60) Roche	CVP >>>	23	8,7	Negative				
(63) Sysmex	CVP >>>	18	6,8	Negative				
(109) Axis-Shield	CVP >>>	1	0,4	Negative				
(150) Arkay		1	0,4	Not specified				
	CVP >>>	14	5,3	Negative				
(162) Siemens (Atellica)	CVP >>>	8	3	Negative				
(179) Siemens	CVP >>>	39	15	Negative				
(197) Analyticon	CVP >>>	2	0,8	Negative				

Summary statistics - qualitative results**EQA round: DGP1/24 - Urine Strip Tests Analysis**

Setup: groups - R (strip manufacturer)

Deadline: 16.2.2024

Test Sample Group	AV	N _{tot}	N _{rel} [%]	Frequency of the results		Success N _{eva} N _{suc} S _{rel} [%]
				Result		
(201) Iris Diagnostics	CVP >>>	14	5,3	Negative		
(202) Dirui	CVP >>>	63	24	Negative		
(207) 77 Elektronika Kft.	CVP >>>	32	12	Negative		
(999) another manufacturer	CVP >>>	5	1,9	Negative		
(273) pH		292				292 282 97
Sample A		292				292 291 100
(5) Dialab	>	3	1	6,0		
(12) Beckman Coulter	>	2	0,7	5,0		
	CVP >>>	6	2,1	5,5		
	>	1	0,3	6,0		
(46) Erba Lachema	>	23	7,9	5,0		
	CVP >>>	4	1,4	5,5		
	>	17	5,8	6,0		
(49) BioVendor	>	1	0,3	5,0		
	CVP >>>	10	3,4	5,5		
(60) Roche	>	14	4,8	5,0		
	CVP >>>	1	0,3	5,5		
	>	10	3,4	6,0		
(63) Sysmex	CVP >>>	18	6,2	5,5		
(109) Axis-Shield	>	1	0,3	6,0		
(150) Arkray	CVP >>>	15	5,1	5,5		
(161) YD Diagnostics	>	1	0,3	6,0		
(162) Siemens (Atellica)	CVP >>>	8	2,7	5,5		
(179) Siemens	CVP >>>	25	8,6	5,5		
	>	15	5,1	6,0		
(197) Analyticon	>	2	0,7	5,0		
(201) Iris Diagnostics	>	2	0,7	5,0		
	CVP >>>	9	3,1	5,5		
	>	2	0,7	6,0		
		1	0,3	8,5		
(202) Dirui	>	1	0,3	5,0		
	CVP >>>	56	19	5,5		
	>	6	2,1	6,0		
(207) 77 Elektronika Kft.	>	2	0,7	5,0		
	CVP >>>	31	11	5,5		
(999) another manufacturer	>	2	0,7	5,0		
	CVP >>>	2	0,7	5,5		
	>	1	0,3	6,0		
Sample B		292				292 282 97
(5) Dialab	CVP >>>	3	1	8,0		
(12) Beckman Coulter	CVP >>>	7	2,4	8,0		
	>	2	0,7	8,5		
(46) Erba Lachema		1	0,3	6,0		
		3	1	6,5		
	>	12	4,1	7,0		
		6	2,1	7,5		
	CVP >>>	20	6,8	8,0		
	>	1	0,3	8,5		
		1	0,3	9,0		
(49) BioVendor	>	3	1	7,5		
	CVP >>>	6	2,1	8,0		
	>	2	0,7	8,5		
(60) Roche	>	1	0,3	7,0		
	CVP >>>	17	5,8	8,0		
	>	1	0,3	8,5		
		6	2,1	9,0		
(63) Sysmex	>	8	2,7	7,5		
	CVP >>>	9	3,1	8,0		
	>	1	0,3	8,5		
(109) Axis-Shield	>	1	0,3	7,0		
(150) Arkray		1	0,3	6,5		
	>	4	1,4	7,5		
	CVP >>>	10	3,4	8,0		
(161) YD Diagnostics	>	1	0,3	7,5		
(162) Siemens (Atellica)	CVP >>>	6	2,1	8,0		
	>	1	0,3	8,5		
		1	0,3	9,0		
(179) Siemens		1	0,3	6,5		
	>	2	0,7	7,0		

Summary statistics - qualitative results**EQA round: DGP1/24 - Urine Strip Tests Analysis**

Deadline: 16.2.2024

Setup: groups - R (strip manufacturer)

Test Sample Group	AV	N _{tot}	N _{rel} [%]	Frequency of the results			Success		
				Result			N _{eva}	N _{suc}	S _{rel} [%]
(197) Analyticon	CVP >>>	2	0,7	7,5					
(201) Iris Diagnostics	CVP >>>	15	5,1	8,0					
		>	16	5,5	8,5				
		>	4	1,4	9,0				
(202) Dirui	CVP >>>	2	0,7	8,0					
			1	0,3	5,5				
	CVP >>>	6	2,1	8,0					
		>	7	2,4	8,5				
(207) 77 Elektronika Kft.	CVP >>>	21	7,2	8,0					
		>	1	0,3	7,0				
		>	9	3,1	7,5				
	CVP >>>	21	7,2	8,0					
		>	2	0,7	8,5				
(999) another manufacturer	CVP >>>	1	0,3	6,0					
		>	1	0,3	7,0				
		>	2	0,7	7,5				
	CVP >>>	1	0,3	8,0					
(266) Urobilinogen		291					291	289	99
Sample A		291					291	289	99
(5) Dialab	CVP >>>	3	1	Negative, < 17 µmol/L					
(12) Beckman Coulter	CVP >>>	9	3,1	Negative, < 17 µmol/L					
(46) Erba Lachema	CVP >>>	43	15	Negative, < 17 µmol/L					
			1	0,3 1+, 2+					
(49) BioVendor	CVP >>>	11	3,8	Negative, < 17 µmol/L					
(60) Roche	CVP >>>	24	8,2	Negative, < 17 µmol/L					
(63) Sysmex	CVP >>>	18	6,2	Negative, < 17 µmol/L					
(109) Axis-Shield	CVP >>>	1	0,3	Negative, < 17 µmol/L					
(150) Arkray	CVP >>>	15	5,2	Negative, < 17 µmol/L					
(161) YD Diagnostics	CVP >>>	1	0,3	Negative, < 17 µmol/L					
(162) Siemens (Atellica)	CVP >>>	8	2,7	Negative, < 17 µmol/L					
(179) Siemens	CVP >>>	40	14	Negative, < 17 µmol/L					
(197) Analyticon	CVP >>>	2	0,7	Negative, < 17 µmol/L					
(201) Iris Diagnostics	CVP >>>	14	4,8	Negative, < 17 µmol/L					
(202) Dirui	CVP >>>	63	22	Negative, < 17 µmol/L					
(207) 77 Elektronika Kft.	CVP >>>	33	11	Negative, < 17 µmol/L					
(999) another manufacturer	CVP >>>	4	1,4	Negative, < 17 µmol/L					
			1	0,3 1+, 2+					
Sample B		291					291	290	100
(5) Dialab	CVP >>>	3	1	Negative, < 17 µmol/L					
(12) Beckman Coulter	CVP >>>	9	3,1	Negative, < 17 µmol/L					
(46) Erba Lachema	CVP >>>	43	15	Negative, < 17 µmol/L					
			1	0,3 1+, 2+					
(49) BioVendor	CVP >>>	11	3,8	Negative, < 17 µmol/L					
(60) Roche	CVP >>>	24	8,2	Negative, < 17 µmol/L					
(63) Sysmex	CVP >>>	18	6,2	Negative, < 17 µmol/L					
(109) Axis-Shield	CVP >>>	1	0,3	Negative, < 17 µmol/L					
(150) Arkray	CVP >>>	15	5,2	Negative, < 17 µmol/L					
(161) YD Diagnostics	CVP >>>	1	0,3	Negative, < 17 µmol/L					
(162) Siemens (Atellica)	CVP >>>	8	2,7	Negative, < 17 µmol/L					
(179) Siemens	CVP >>>	40	14	Negative, < 17 µmol/L					
(197) Analyticon	CVP >>>	2	0,7	Negative, < 17 µmol/L					
(201) Iris Diagnostics	CVP >>>	14	4,8	Negative, < 17 µmol/L					
(202) Dirui	CVP >>>	63	22	Negative, < 17 µmol/L					
(207) 77 Elektronika Kft.	CVP >>>	33	11	Negative, < 17 µmol/L					
(999) another manufacturer	CVP >>>	5	1,7	Negative, < 17 µmol/L					