

Summary statistics - qualitative results**EQA round: DGP1/24 - Urine Strip Tests Analysis**

Deadline: 16.2.2024

Setup: groups - S (measuring system)

AV = assigned value CVP = consensus of all participants >>> ... expected result > ... acceptable result ± ... result not assessed	N _{tot} = total number of the results N _{rel} = relative number of the results	N _{eva} = number of the results assessed N _{suc} = number of successful results S _{rel} = relative success
---	---	---

Test Sample Group	AV	Frequency of the results			Success N _{eva} N _{suc} S _{rel} [%]
		N _{tot}	N _{rel}	Result [%]	
(268) Bilirubin		291			291 289 99
Sample A		291			291 289 99
(201) Arkray Aution JET / Eleven	CVP >>>	5	1,7	Negative	
(202) Arkray Aution MAX / Sysmex UX-2000	CVP >>>	10	3,4	Negative	
(204) Siemens Clinitek Novus	CVP >>>	34	12	Negative	
(205) Roche cobas u 601 (cobas 6500)	CVP >>>	9	3,1	Negative	
(208) Siemens Clinitek Status	CVP >>>	4	1,4	Negative	
(209) 77 Elektronika LabUreader	CVP >>>	10	3,4	Negative	
(215) Roche Urisys 1800 / cobas u411	CVP >>>	13	4,5	Negative	
(216) Sysmex UC-1000, UC-3500	CVP >>>	17	5,8	Negative	
		1	0,3	1+, 2+	
(218) YD Diagnostics URISCAN Pro+	CVP >>>	1	0,3	Negative	
(220) 77 Elektronika HandUReader	CVP >>>	1	0,3	Negative	
(221) Erba Lachema LAURA	CVP >>>	18	6,2	Negative	
(222) Roche Urisys 1100	CVP >>>	3	1	Negative	
(223) Iris iChem Velocity	CVP >>>	23	7,9	Negative	
(225) Dirui H-series	CVP >>>	10	3,4	Negative	
(228) Siemens Clinitek Advantus	CVP >>>	10	3,4	Negative	
(229) 77 Elektronika LabUMat, LabUMat 2	CVP >>>	31	11	Negative	
(230) Dirui FUS-series	CVP >>>	52	18	Negative	
(998) Manual	CVP >>>	24	8,2	Negative	
		1	0,3	1+, 2+	
(999) another system	CVP >>>	14	4,8	Negative	
Sample B		291			291 290 100
(201) Arkray Aution JET / Eleven	CVP >>>	5	1,7	Negative	
(202) Arkray Aution MAX / Sysmex UX-2000	CVP >>>	10	3,4	Negative	
(204) Siemens Clinitek Novus	CVP >>>	34	12	Negative	
(205) Roche cobas u 601 (cobas 6500)	CVP >>>	9	3,1	Negative	
(208) Siemens Clinitek Status	CVP >>>	4	1,4	Negative	
(209) 77 Elektronika LabUreader	CVP >>>	10	3,4	Negative	
(215) Roche Urisys 1800 / cobas u411	CVP >>>	13	4,5	Negative	
(216) Sysmex UC-1000, UC-3500	CVP >>>	18	6,2	Negative	
(218) YD Diagnostics URISCAN Pro+	CVP >>>	1	0,3	Negative	
(220) 77 Elektronika HandUReader	CVP >>>	1	0,3	Negative	
(221) Erba Lachema LAURA	CVP >>>	18	6,2	Negative	
(222) Roche Urisys 1100	CVP >>>	3	1	Negative	
(223) Iris iChem Velocity	CVP >>>	23	7,9	Negative	
(225) Dirui H-series	CVP >>>	10	3,4	Negative	
(228) Siemens Clinitek Advantus	CVP >>>	10	3,4	Negative	
(229) 77 Elektronika LabUMat, LabUMat 2	CVP >>>	31	11	Negative	
(230) Dirui FUS-series	CVP >>>	52	18	Negative	
(998) Manual	CVP >>>	24	8,2	Negative	
		1	0,3	1+, 2+	
(999) another system	CVP >>>	14	4,8	Negative	
(264) Total protein		291			291 284 98
Sample A		291			291 289 99
(201) Arkray Aution JET / Eleven	CVP >>>	5	1,7	Negative, < 0,3 g/L	
(202) Arkray Aution MAX / Sysmex UX-2000	CVP >>>	10	3,4	Negative, < 0,3 g/L	
		1	0,3	0,3 - 0,5 g/L	
(204) Siemens Clinitek Novus	CVP >>>	33	11	Negative, < 0,3 g/L	
(205) Roche cobas u 601 (cobas 6500)	CVP >>>	9	3,1	Negative, < 0,3 g/L	
(208) Siemens Clinitek Status	CVP >>>	4	1,4	Negative, < 0,3 g/L	
(209) 77 Elektronika LabUreader	CVP >>>	10	3,4	Negative, < 0,3 g/L	
(215) Roche Urisys 1800 / cobas u411	CVP >>>	13	4,5	Negative, < 0,3 g/L	
(216) Sysmex UC-1000, UC-3500	CVP >>>	18	6,2	Negative, < 0,3 g/L	
(218) YD Diagnostics URISCAN Pro+	CVP >>>	1	0,3	Negative, < 0,3 g/L	
(220) 77 Elektronika HandUReader	CVP >>>	1	0,3	Negative, < 0,3 g/L	
(221) Erba Lachema LAURA	CVP >>>	18	6,2	Negative, < 0,3 g/L	
(222) Roche Urisys 1100	CVP >>>	3	1	Negative, < 0,3 g/L	
(223) Iris iChem Velocity	CVP >>>	22	7,6	Negative, < 0,3 g/L	
		1	0,3	0,3 - 0,5 g/L	
(225) Dirui H-series	CVP >>>	9	3,1	Negative, < 0,3 g/L	
(228) Siemens Clinitek Advantus	CVP >>>	10	3,4	Negative, < 0,3 g/L	
(229) 77 Elektronika LabUMat, LabUMat 2	CVP >>>	31	11	Negative, < 0,3 g/L	

Summary statistics - qualitative results**EQA round: DGP1/24 - Urine Strip Tests Analysis**

Deadline: 16.2.2024

Setup: groups - S (measuring system)

Test Sample Group	AV	N _{tot}	N _{rel} [%]	Frequency of the results		N _{eva}	N _{suc}	S _{rel} [%]
				Result				
(230) Dirui FUS-series	CVP >>>	53	18	Negative, < 0,3 g/L				
(998) Manual	CVP >>>	26	8,9	Negative, < 0,3 g/L				
(999) another system	CVP >>>	13	4,5	Negative, < 0,3 g/L				
Sample B		291				291	286	98
(201) Arkray Aution JET / Eleven	CVP >>>	4	1,4	Negative, < 0,3 g/L				
	CVP >>>	1	0,3	0,3 - 0,5 g/L				
(202) Arkray Aution MAX / Sysmex UX-2000	CVP >>>	6	2,1	Negative, < 0,3 g/L				
	CVP >>>	5	1,7	0,3 - 0,5 g/L				
(204) Siemens Clinitek Novus	CVP >>>	7	2,4	Negative, < 0,3 g/L				
	CVP >>>	26	8,9	0,3 - 0,5 g/L				
(205) Roche cobas u 601 (cobas 6500)	CVP >>>	2	0,7	Negative, < 0,3 g/L				
	CVP >>>	7	2,4	0,3 - 0,5 g/L				
(208) Siemens Clinitek Status	CVP >>>	1	0,3	Negative, < 0,3 g/L				
	CVP >>>	2	0,7	0,3 - 0,5 g/L				
		1	0,3	0,51 - 2,5 g/L				
(209) 77 Elektronika LabUreader	CVP >>>	8	2,7	Negative, < 0,3 g/L				
	CVP >>>	2	0,7	0,3 - 0,5 g/L				
(215) Roche Urisys 1800 / cobas u411	CVP >>>	5	1,7	Negative, < 0,3 g/L				
	CVP >>>	8	2,7	0,3 - 0,5 g/L				
(216) Sysmex UC-1000, UC-3500	CVP >>>	10	3,4	Negative, < 0,3 g/L				
	CVP >>>	8	2,7	0,3 - 0,5 g/L				
(218) YD Diagnostics URISCAN Pro+	CVP >>>	1	0,3	0,3 - 0,5 g/L				
(220) 77 Elektronika HandUReader	CVP >>>	1	0,3	0,3 - 0,5 g/L				
(221) Erba Lachema LAURA	CVP >>>	10	3,4	Negative, < 0,3 g/L				
	CVP >>>	8	2,7	0,3 - 0,5 g/L				
(222) Roche Urisys 1100	CVP >>>	3	1	Negative, < 0,3 g/L				
(223) Iris iChem Velocity	CVP >>>	5	1,7	Negative, < 0,3 g/L				
	CVP >>>	17	5,8	0,3 - 0,5 g/L				
		1	0,3	0,51 - 2,5 g/L				
(225) Dirui H-series	CVP >>>	2	0,7	Negative, < 0,3 g/L				
	CVP >>>	7	2,4	0,3 - 0,5 g/L				
(228) Siemens Clinitek Advantus	CVP >>>	7	2,4	Negative, < 0,3 g/L				
	CVP >>>	3	1	0,3 - 0,5 g/L				
(229) 77 Elektronika LabUMat, LabUMat 2	CVP >>>	19	6,5	Negative, < 0,3 g/L				
	CVP >>>	12	4,1	0,3 - 0,5 g/L				
(230) Dirui FUS-series	CVP >>>	21	7,2	Negative, < 0,3 g/L				
	CVP >>>	30	10	0,3 - 0,5 g/L				
		2	0,7	0,51 - 2,5 g/L				
(998) Manual	CVP >>>	9	3,1	Negative, < 0,3 g/L				
	CVP >>>	16	5,5	0,3 - 0,5 g/L				
		1	0,3	0,51 - 2,5 g/L				
(999) another system	CVP >>>	9	3,1	Negative, < 0,3 g/L				
	CVP >>>	4	1,4	0,3 - 0,5 g/L				
(261) Glucose		292				292	265	91
Sample A		292				292	265	91
(201) Arkray Aution JET / Eleven	CVP >>>	3	1	2,81 - 11 mmol/L				
	CVP >>>	2	0,7	11,1 - 28 mmol/L				
(202) Arkray Aution MAX / Sysmex UX-2000	CVP >>>	4	1,4	2,81 - 11 mmol/L				
	CVP >>>	6	2,1	11,1 - 28 mmol/L				
(204) Siemens Clinitek Novus		2	0,7	Negative, normal, <= 2,8 mmol/L				
	CVP >>>	16	5,5	2,81 - 11 mmol/L				
	CVP >>>	16	5,5	11,1 - 28 mmol/L				
(205) Roche cobas u 601 (cobas 6500)	CVP >>>	1	0,3	2,81 - 11 mmol/L				
	CVP >>>	8	2,7	11,1 - 28 mmol/L				
(208) Siemens Clinitek Status	CVP >>>	1	0,3	2,81 - 11 mmol/L				
	CVP >>>	3	1	11,1 - 28 mmol/L				
(209) 77 Elektronika LabUreader	CVP >>>	5	1,7	2,81 - 11 mmol/L				
	CVP >>>	3	1	11,1 - 28 mmol/L				
		2	0,7	> 28 mmol/L				
(215) Roche Urisys 1800 / cobas u411	CVP >>>	9	3,1	2,81 - 11 mmol/L				
	CVP >>>	4	1,4	11,1 - 28 mmol/L				
(216) Sysmex UC-1000, UC-3500	CVP >>>	3	1	2,81 - 11 mmol/L				
	CVP >>>	15	5,1	11,1 - 28 mmol/L				
(218) YD Diagnostics URISCAN Pro+	CVP >>>	1	0,3	11,1 - 28 mmol/L				
(220) 77 Elektronika HandUReader		1	0,3	> 28 mmol/L				
(221) Erba Lachema LAURA	CVP >>>	2	0,7	2,81 - 11 mmol/L				
	CVP >>>	9	3,1	11,1 - 28 mmol/L				
		7	2,4	> 28 mmol/L				
(222) Roche Urisys 1100	CVP >>>	1	0,3	2,81 - 11 mmol/L				

Summary statistics - qualitative results**EQA round: DGP1/24 - Urine Strip Tests Analysis**

Deadline: 16.2.2024

Setup: groups - S (measuring system)

Test Sample Group	AV	N _{tot}	N _{rel} [%]	Frequency of the results		N _{eva}	N _{suc}	S _{rel} [%]
				Result				
(223) Iris iChem Velocity	CVP >>>	2	0,7	11,1 - 28 mmol/L				
		2	0,7	Negative, normal, <= 2,8 mmol/L				
(225) Dirui H-series	CVP >>>	17	5,8	2,81 - 11 mmol/L				
	CVP >>>	3	1	11,1 - 28 mmol/L				
		1	0,3	> 28 mmol/L				
(228) Siemens Clinitek Advantus	CVP >>>	3	1	2,81 - 11 mmol/L				
	CVP >>>	6	2,1	11,1 - 28 mmol/L				
(229) 77 Elektronika LabUMat, LabUMat 2	CVP >>>	1	0,3	2,81 - 11 mmol/L				
	CVP >>>	9	3,1	11,1 - 28 mmol/L				
	CVP >>>	18	6,2	2,81 - 11 mmol/L				
		10	3,4	11,1 - 28 mmol/L				
		3	1	> 28 mmol/L				
(230) Dirui FUS-series		1	0,3	Negative, normal, <= 2,8 mmol/L				
	CVP >>>	21	7,2	2,81 - 11 mmol/L				
	CVP >>>	30	10	11,1 - 28 mmol/L				
		1	0,3	> 28 mmol/L				
(998) Manual	CVP >>>	7	2,4	2,81 - 11 mmol/L				
	CVP >>>	13	4,5	11,1 - 28 mmol/L				
		6	2,1	> 28 mmol/L				
(999) another system	CVP >>>	6	2,1	2,81 - 11 mmol/L				
	CVP >>>	7	2,4	11,1 - 28 mmol/L				
		1	0,3	> 28 mmol/L				
Sample B		292				292	289	99
(201) Arkray Aution JET / Eleven	CVP >>>	5	1,7	Negative, normal, <= 2,8 mmol/L				
(202) Arkray Aution MAX / Sysmex UX-2000	CVP >>>	10	3,4	Negative, normal, <= 2,8 mmol/L				
(204) Siemens Clinitek Novus	CVP >>>	33	11	Negative, normal, <= 2,8 mmol/L				
		1	0,3	2,81 - 11 mmol/L				
(205) Roche cobas u 601 (cobas 6500)	CVP >>>	9	3,1	Negative, normal, <= 2,8 mmol/L				
(208) Siemens Clinitek Status	CVP >>>	4	1,4	Negative, normal, <= 2,8 mmol/L				
(209) 77 Elektronika LabUreader	CVP >>>	10	3,4	Negative, normal, <= 2,8 mmol/L				
(215) Roche Urisys 1800 / cobas u411	CVP >>>	13	4,5	Negative, normal, <= 2,8 mmol/L				
(216) Sysmex UC-1000, UC-3500	CVP >>>	18	6,2	Negative, normal, <= 2,8 mmol/L				
(218) YD Diagnostics URISCAN Pro+	CVP >>>	1	0,3	Negative, normal, <= 2,8 mmol/L				
(220) 77 Elektronika HandUReader	CVP >>>	1	0,3	Negative, normal, <= 2,8 mmol/L				
(221) Erba Lachema LAURA	CVP >>>	18	6,2	Negative, normal, <= 2,8 mmol/L				
(222) Roche Urisys 1100	CVP >>>	3	1	Negative, normal, <= 2,8 mmol/L				
(223) Iris iChem Velocity	CVP >>>	22	7,5	Negative, normal, <= 2,8 mmol/L				
		1	0,3	2,81 - 11 mmol/L				
(225) Dirui H-series	CVP >>>	9	3,1	Negative, normal, <= 2,8 mmol/L				
(228) Siemens Clinitek Advantus	CVP >>>	10	3,4	Negative, normal, <= 2,8 mmol/L				
(229) 77 Elektronika LabUMat, LabUMat 2	CVP >>>	31	11	Negative, normal, <= 2,8 mmol/L				
(230) Dirui FUS-series	CVP >>>	53	18	Negative, normal, <= 2,8 mmol/L				
(998) Manual	CVP >>>	26	8,9	Negative, normal, <= 2,8 mmol/L				
(999) another system	CVP >>>	13	4,5	Negative, normal, <= 2,8 mmol/L				
		1	0,3	2,81 - 11 mmol/L				
(275) hCG		11				11	11	100
Sample A		11				11	11	100
(208) Siemens Clinitek Status	CVP >>>	1	9,1	Negative				
(998) Manual	CVP >>>	10	91	Negative				
Sample B		11				11	11	100
(208) Siemens Clinitek Status	CVP >>>	1	9,1	Positive				
(998) Manual	CVP >>>	10	91	Positive				
(263) Ketone bodies		291				291	290	100
Sample A		291				291	290	100
(201) Arkray Aution JET / Eleven	CVP >>>	5	1,7	Negative, <= 0,5 mmol/L				
(202) Arkray Aution MAX / Sysmex UX-2000	CVP >>>	10	3,4	Negative, <= 0,5 mmol/L				
(204) Siemens Clinitek Novus	CVP >>>	34	12	Negative, <= 0,5 mmol/L				
(205) Roche cobas u 601 (cobas 6500)	CVP >>>	9	3,1	Negative, <= 0,5 mmol/L				
(208) Siemens Clinitek Status	CVP >>>	4	1,4	Negative, <= 0,5 mmol/L				
(209) 77 Elektronika LabUreader	CVP >>>	10	3,4	Negative, <= 0,5 mmol/L				
(215) Roche Urisys 1800 / cobas u411	CVP >>>	13	4,5	Negative, <= 0,5 mmol/L				
(216) Sysmex UC-1000, UC-3500	CVP >>>	18	6,2	Negative, <= 0,5 mmol/L				
(218) YD Diagnostics URISCAN Pro+	CVP >>>	1	0,3	Negative, <= 0,5 mmol/L				
(220) 77 Elektronika HandUReader	CVP >>>	1	0,3	Negative, <= 0,5 mmol/L				
(221) Erba Lachema LAURA	CVP >>>	18	6,2	Negative, <= 0,5 mmol/L				
(222) Roche Urisys 1100	CVP >>>	3	1	Negative, <= 0,5 mmol/L				
(223) Iris iChem Velocity	CVP >>>	23	7,9	Negative, <= 0,5 mmol/L				
(225) Dirui H-series	CVP >>>	10	3,4	Negative, <= 0,5 mmol/L				
(228) Siemens Clinitek Advantus	CVP >>>	9	3,1	Negative, <= 0,5 mmol/L				

Summary statistics - qualitative results**EQA round: DGP1/24 - Urine Strip Tests Analysis**

Deadline: 16.2.2024

Setup: groups - S (measuring system)

Test Sample Group	AV	N _{tot}	N _{rel} [%]	Frequency of the results		N _{eva}	N _{suc}	S _{rel} [%]
				Result				
(229) 77 Elektronika LabUMat, LabUMat 2	CVP >>>	31	11	Negative, <= 0,5 mmol/L				
(230) Dirui FUS-series	CVP >>>	52	18	Negative, <= 0,5 mmol/L				
(998) Manual	CVP >>>	25	8,6	Negative, <= 0,5 mmol/L				
		1	0,3	1+, 2+				
(999) another system	CVP >>>	14	4,8	Negative, <= 0,5 mmol/L				
Sample B		291				291	290	100
(201) ArkRAY Aution JET / Eleven	CVP >>>	5	1,7	Negative, <= 0,5 mmol/L				
(202) ArkRAY Aution MAX / Sysmex UX-2000	CVP >>>	10	3,4	Negative, <= 0,5 mmol/L				
(204) Siemens Clinitek Novus	CVP >>>	34	12	Negative, <= 0,5 mmol/L				
(205) Roche cobas u 601 (cobas 6500)	CVP >>>	9	3,1	Negative, <= 0,5 mmol/L				
(208) Siemens Clinitek Status	CVP >>>	4	1,4	Negative, <= 0,5 mmol/L				
(209) 77 Elektronika LabUReader	CVP >>>	10	3,4	Negative, <= 0,5 mmol/L				
(215) Roche Urisys 1800 / cobas u411	CVP >>>	13	4,5	Negative, <= 0,5 mmol/L				
(216) Sysmex UC-1000, UC-3500	CVP >>>	18	6,2	Negative, <= 0,5 mmol/L				
(218) YD Diagnostics URISCAN Pro+	CVP >>>	1	0,3	Negative, <= 0,5 mmol/L				
(220) 77 Elektronika HandUReader	CVP >>>	1	0,3	Negative, <= 0,5 mmol/L				
(221) Erba Lachema LAURA	CVP >>>	18	6,2	Negative, <= 0,5 mmol/L				
(222) Roche Urisys 1100	CVP >>>	3	1	Negative, <= 0,5 mmol/L				
(223) Iris iChem Velocity	CVP >>>	23	7,9	Negative, <= 0,5 mmol/L				
(225) Dirui H-series	CVP >>>	10	3,4	Negative, <= 0,5 mmol/L				
(228) Siemens Clinitek Advantus	CVP >>>	9	3,1	Negative, <= 0,5 mmol/L				
(229) 77 Elektronika LabUMat, LabUMat 2	CVP >>>	31	11	Negative, <= 0,5 mmol/L				
(230) Dirui FUS-series	CVP >>>	52	18	Negative, <= 0,5 mmol/L				
(998) Manual	CVP >>>	25	8,6	Negative, <= 0,5 mmol/L				
		1	0,3	1+, 2+				
(999) another system	CVP >>>	14	4,8	Negative, <= 0,5 mmol/L				
(272) Blood		292				292	280	96
Sample A		292				292	285	98
(201) ArkRAY Aution JET / Eleven	CVP >>>	5	1,7	> 50 ery/ μ L				
(202) ArkRAY Aution MAX / Sysmex UX-2000		1	0,3	5 - 12 ery/ μ L				
	>	2	0,7	13 - 50 ery/ μ L				
	CVP >>>	7	2,4	> 50 ery/ μ L				
(204) Siemens Clinitek Novus		>	7	2,4	13 - 50 ery/ μ L			
	CVP >>>	27	9,2	> 50 ery/ μ L				
(205) Roche cobas u 601 (cobas 6500)		>	1	0,3	13 - 50 ery/ μ L			
	CVP >>>	8	2,7	> 50 ery/ μ L				
(208) Siemens Clinitek Status	CVP >>>	4	1,4	> 50 ery/ μ L				
(209) 77 Elektronika LabUReader	CVP >>>	10	3,4	> 50 ery/ μ L				
(215) Roche Urisys 1800 / cobas u411		1	0,3	5 - 12 ery/ μ L				
	CVP >>>	12	4,1	> 50 ery/ μ L				
(216) Sysmex UC-1000, UC-3500		>	1	0,3	13 - 50 ery/ μ L			
	CVP >>>	17	5,8	> 50 ery/ μ L				
(218) YD Diagnostics URISCAN Pro+	CVP >>>	1	0,3	> 50 ery/ μ L				
(220) 77 Elektronika HandUReader	CVP >>>	1	0,3	> 50 ery/ μ L				
(221) Erba Lachema LAURA		1	0,3	5 - 12 ery/ μ L				
	>	6	2,1	13 - 50 ery/ μ L				
	CVP >>>	11	3,8	> 50 ery/ μ L				
(222) Roche Urisys 1100	CVP >>>	3	1	> 50 ery/ μ L				
(223) Iris iChem Velocity		2	0,7	5 - 12 ery/ μ L				
	>	3	1	13 - 50 ery/ μ L				
	CVP >>>	18	6,2	> 50 ery/ μ L				
(225) Dirui H-series	CVP >>>	9	3,1	> 50 ery/ μ L				
(228) Siemens Clinitek Advantus	CVP >>>	10	3,4	> 50 ery/ μ L				
(229) 77 Elektronika LabUMat, LabUMat 2	CVP >>>	31	11	> 50 ery/ μ L				
(230) Dirui FUS-series		>	6	2,1	13 - 50 ery/ μ L			
	CVP >>>	47	16	> 50 ery/ μ L				
(998) Manual		2	0,7	5 - 12 ery/ μ L				
	>	8	2,7	13 - 50 ery/ μ L				
	CVP >>>	16	5,5	> 50 ery/ μ L				
(999) another system		>	3	1	13 - 50 ery/ μ L			
	CVP >>>	11	3,8	> 50 ery/ μ L				
Sample B		292				292	286	98
(201) ArkRAY Aution JET / Eleven	CVP >>>	3	1	Negative				
	>	2	0,7	5 - 12 ery/ μ L				
(202) ArkRAY Aution MAX / Sysmex UX-2000	CVP >>>	10	3,4	Negative				
(204) Siemens Clinitek Novus	CVP >>>	33	11	Negative				
	>	1	0,3	5 - 12 ery/ μ L				
(205) Roche cobas u 601 (cobas 6500)	CVP >>>	8	2,7	Negative				
	>	1	0,3	5 - 12 ery/ μ L				

Summary statistics - qualitative results**EQA round: DGP1/24 - Urine Strip Tests Analysis**

Deadline: 16.2.2024

Setup: groups - S (measuring system)

Test Sample Group	AV	N _{tot}	N _{rel} [%]	Frequency of the results		Success N _{eva} N _{suc} S _{rel} [%]
				Result		
(208) Siemens Clinitek Status	CVP >>>	2	0,7	Negative		
		>	2	0,7	5 - 12 ery/µL	
(209) 77 Elektronika LabUreader	CVP >>>	4	1,4	Negative		
		>	6	2,1	5 - 12 ery/µL	
(215) Roche Urisys 1800 / cobas u411	CVP >>>	12	4,1	Negative		
		>	1	0,3	5 - 12 ery/µL	
(216) Sysmex UC-1000, UC-3500	CVP >>>	18	6,2	Negative		
(218) YD Diagnostics URISCAN Pro+		>	1	0,3	5 - 12 ery/µL	
(220) 77 Elektronika HandUReader	CVP >>>	1	0,3	Negative		
(221) Erba Lachema LAURA	CVP >>>	18	6,2	Negative		
(222) Roche Urisys 1100	CVP >>>	2	0,7	Negative		
		>	1	0,3	5 - 12 ery/µL	
(223) Iris iChem Velocity	CVP >>>	8	2,7	Negative		
		>	13	4,5	5 - 12 ery/µL	
			1	0,3	13 - 50 ery/µL	
			1	0,3	> 50 ery/µL	
(225) Dirui H-series	CVP >>>	9	3,1	Negative		
(228) Siemens Clinitek Advantus	CVP >>>	4	1,4	Negative		
		>	5	1,7	5 - 12 ery/µL	
			1	0,3	13 - 50 ery/µL	
(229) 77 Elektronika LabUMat, LabUMat 2	CVP >>>	8	2,7	Negative		
		>	22	7,5	5 - 12 ery/µL	
			1	0,3	13 - 50 ery/µL	
(230) Dirui FUS-series	CVP >>>	37	13	Negative		
		>	15	5,1	5 - 12 ery/µL	
			1	0,3	> 50 ery/µL	
(998) Manual	CVP >>>	12	4,1	Negative		
		>	14	4,8	5 - 12 ery/µL	
(999) another system	CVP >>>	9	3,1	Negative		
		>	4	1,4	5 - 12 ery/µL	
			1	0,3	13 - 50 ery/µL	
(270) WBC		262				262 260 99
Sample A		262				262 261 100
(0) not specified	CVP >>>	2	0,8	Negative		
(201) Arkray Auton JET / Eleven	CVP >>>	5	1,9	Negative		
(202) Arkray Auton MAX / Sysmex UX-2000	CVP >>>	10	3,8	Negative		
(204) Siemens Clinitek Novus	CVP >>>	32	12	Negative		
(205) Roche cobas u 601 (cobas 6500)	CVP >>>	9	3,4	Negative		
(208) Siemens Clinitek Status	CVP >>>	4	1,5	Negative		
(209) 77 Elektronika LabUreader	CVP >>>	10	3,8	Negative		
(215) Roche Urisys 1800 / cobas u411	CVP >>>	13	5	Negative		
(216) Sysmex UC-1000, UC-3500	CVP >>>	18	6,9	Negative		
(220) 77 Elektronika HandUReader	CVP >>>	1	0,4	Negative		
(221) Erba Lachema LAURA	CVP >>>	15	5,7	Negative		
(222) Roche Urisys 1100	CVP >>>	2	0,8	Negative		
(223) Iris iChem Velocity		1	0,4	Not specified		
	CVP >>>	20	7,6	Negative		
(225) Dirui H-series	CVP >>>	9	3,4	Negative		
(228) Siemens Clinitek Advantus	CVP >>>	11	4,2	Negative		
(229) 77 Elektronika LabUMat, LabUMat 2	CVP >>>	31	12	Negative		
(230) Dirui FUS-series	CVP >>>	53	20	Negative		
(998) Manual	CVP >>>	3	1,1	Negative		
(999) another system	CVP >>>	13	5	Negative		
Sample B		262				262 261 100
(0) not specified	CVP >>>	2	0,8	Negative		
(201) Arkray Auton JET / Eleven	CVP >>>	5	1,9	Negative		
(202) Arkray Auton MAX / Sysmex UX-2000	CVP >>>	10	3,8	Negative		
(204) Siemens Clinitek Novus	CVP >>>	32	12	Negative		
(205) Roche cobas u 601 (cobas 6500)	CVP >>>	9	3,4	Negative		
(208) Siemens Clinitek Status	CVP >>>	4	1,5	Negative		
(209) 77 Elektronika LabUreader	CVP >>>	10	3,8	Negative		
(215) Roche Urisys 1800 / cobas u411	CVP >>>	13	5	Negative		
(216) Sysmex UC-1000, UC-3500	CVP >>>	18	6,9	Negative		
(220) 77 Elektronika HandUReader	CVP >>>	1	0,4	Negative		
(221) Erba Lachema LAURA	CVP >>>	15	5,7	Negative		
(222) Roche Urisys 1100	CVP >>>	2	0,8	Negative		
(223) Iris iChem Velocity	CVP >>>	21	8	Negative		
(225) Dirui H-series	CVP >>>	9	3,4	Negative		
(228) Siemens Clinitek Advantus	CVP >>>	11	4,2	Negative		

Summary statistics - qualitative results**EQA round: DGP1/24 - Urine Strip Tests Analysis**

Deadline: 16.2.2024

Setup: groups - S (measuring system)

Test Sample Group	AV	N _{tot}	N _{rel} [%]	Frequency of the results		N _{eva}	N _{suc}	S _{rel} [%]
				Result				
(229) 77 Elektronika LabUMat, LabUMat 2	CVP >>>	31	12	Negative				
(230) Dirui FUS-series	CVP >>>	53	20	Negative				
(998) Manual	CVP >>>	3	1,1	Negative				
(999) another system	CVP >>>	12	4,6	Negative				
		1	0,4	10 - 50 leu/ μ L				
(274) Nitrite		263				263	262	100
Sample A		263				263	263	100
(201) ArkRAY Autio JET / Eleven	CVP >>>	5	1,9	Negative				
(202) ArkRAY Autio MAX / Sysmex UX-2000	CVP >>>	10	3,8	Negative				
(204) Siemens Clinitek Novus	CVP >>>	34	13	Negative				
(205) Roche cobas u 601 (cobas 6500)	CVP >>>	9	3,4	Negative				
(208) Siemens Clinitek Status	CVP >>>	4	1,5	Negative				
(209) 77 Elektronika LabUReader	CVP >>>	10	3,8	Negative				
(215) Roche Urisys 1800 / cobas u411	CVP >>>	13	4,9	Negative				
(216) Sysmex UC-1000, UC-3500	CVP >>>	18	6,8	Negative				
(220) 77 Elektronika HandUReader	CVP >>>	1	0,4	Negative				
(221) Erba Lachema LAURA	CVP >>>	15	5,7	Negative				
(222) Roche Urisys 1100	CVP >>>	2	0,8	Negative				
(223) Iris iChem Velocity	CVP >>>	23	8,7	Negative				
(225) Dirui H-series	CVP >>>	9	3,4	Negative				
(228) Siemens Clinitek Advantus	CVP >>>	10	3,8	Negative				
(229) 77 Elektronika LabUMat, LabUMat 2	CVP >>>	30	11	Negative				
(230) Dirui FUS-series	CVP >>>	52	20	Negative				
(998) Manual	CVP >>>	4	1,5	Negative				
(999) another system	CVP >>>	14	5,3	Negative				
Sample B		263				263	262	100
(201) ArkRAY Autio JET / Eleven		1	0,4	Not specified				
	CVP >>>	4	1,5	Negative				
(202) ArkRAY Autio MAX / Sysmex UX-2000	CVP >>>	10	3,8	Negative				
(204) Siemens Clinitek Novus	CVP >>>	34	13	Negative				
(205) Roche cobas u 601 (cobas 6500)	CVP >>>	9	3,4	Negative				
(208) Siemens Clinitek Status	CVP >>>	4	1,5	Negative				
(209) 77 Elektronika LabUReader	CVP >>>	10	3,8	Negative				
(215) Roche Urisys 1800 / cobas u411	CVP >>>	13	4,9	Negative				
(216) Sysmex UC-1000, UC-3500	CVP >>>	18	6,8	Negative				
(220) 77 Elektronika HandUReader	CVP >>>	1	0,4	Negative				
(221) Erba Lachema LAURA	CVP >>>	15	5,7	Negative				
(222) Roche Urisys 1100	CVP >>>	2	0,8	Negative				
(223) Iris iChem Velocity	CVP >>>	23	8,7	Negative				
(225) Dirui H-series	CVP >>>	9	3,4	Negative				
(228) Siemens Clinitek Advantus	CVP >>>	10	3,8	Negative				
(229) 77 Elektronika LabUMat, LabUMat 2	CVP >>>	30	11	Negative				
(230) Dirui FUS-series	CVP >>>	52	20	Negative				
(998) Manual	CVP >>>	4	1,5	Negative				
(999) another system	CVP >>>	14	5,3	Negative				
(273) pH		292				292	282	97
Sample A		292				292	291	100
(201) ArkRAY Autio JET / Eleven	CVP >>>	5	1,7	5,5				
(202) ArkRAY Autio MAX / Sysmex UX-2000	CVP >>>	11	3,8	5,5				
(204) Siemens Clinitek Novus	CVP >>>	28	9,6	5,5				
	>	6	2,1	6,0				
(205) Roche cobas u 601 (cobas 6500)	>	1	0,3	5,0				
	>	8	2,7	6,0				
(208) Siemens Clinitek Status	CVP >>>	2	0,7	5,5				
	>	2	0,7	6,0				
(209) 77 Elektronika LabUReader	>	4	1,4	5,0				
	CVP >>>	6	2,1	5,5				
(215) Roche Urisys 1800 / cobas u411	>	10	3,4	5,0				
	CVP >>>	1	0,3	5,5				
	>	2	0,7	6,0				
(216) Sysmex UC-1000, UC-3500	CVP >>>	17	5,8	5,5				
(218) YD Diagnostics URISCAN Pro+	>	1	0,3	6,0				
(220) 77 Elektronika HandUReader	CVP >>>	1	0,3	5,5				
(221) Erba Lachema LAURA	>	10	3,4	5,0				
	>	8	2,7	6,0				
(222) Roche Urisys 1100	>	3	1	5,0				
(223) Iris iChem Velocity	>	4	1,4	5,0				
	CVP >>>	15	5,1	5,5				
	>	3	1	6,0				

Summary statistics - qualitative results**EQA round: DGP1/24 - Urine Strip Tests Analysis**

Deadline: 16.2.2024

Setup: groups - S (measuring system)

Test Sample Group	AV	N _{tot}	N _{rel} [%]	Frequency of the results		Success N _{eva}	N _{suc}	S _{rel} [%]
				Result				
(225) Dirui H-series		1	0,3	8,5				
	>	1	0,3	5,0				
	CVP	>>>	7	2,4	5,5			
		>	1	0,3	6,0			
(228) Siemens Clinitek Advantus	CVP	>>>	2	0,7	5,5			
		>	8	2,7	6,0			
(229) 77 Elektronika LabUMat, LabUMat 2	CVP	>>>	31	11	5,5			
(230) Dirui FUS-series	CVP	>>>	49	17	5,5			
		>	4	1,4	6,0			
(998) Manual		>	13	4,5	5,0			
	CVP	>>>	4	1,4	5,5			
		>	9	3,1	6,0			
(999) another system		>	3	1	5,0			
	CVP	>>>	6	2,1	5,5			
		>	5	1,7	6,0			
Sample B		292				292	282	97
(201) Arkray Aution JET / Eleven		1	0,3	6,5				
	>	3	1	7,5				
	CVP	>>>	1	0,3	8,0			
(202) Arkray Aution MAX / Sysmex UX-2000		>	2	0,7	7,5			
	CVP	>>>	9	3,1	8,0			
(204) Siemens Clinitek Novus	CVP	>>>	15	5,1	8,0			
		>	14	4,8	8,5			
		>	5	1,7	9,0			
(205) Roche cobas u 601 (cobas 6500)	CVP	>>>	6	2,1	8,0			
		>	1	0,3	8,5			
		>	2	0,7	9,0			
(208) Siemens Clinitek Status		>	1	0,3	7,5			
	CVP	>>>	1	0,3	8,0			
		>	2	0,7	8,5			
(209) 77 Elektronika LabUreader		>	1	0,3	7,0			
		>	3	1	7,5			
	CVP	>>>	6	2,1	8,0			
(215) Roche Urisys 1800 / cobas u411	CVP	>>>	9	3,1	8,0			
		>	4	1,4	9,0			
(216) Sysmex UC-1000, UC-3500		>	7	2,4	7,5			
	CVP	>>>	9	3,1	8,0			
		>	1	0,3	8,5			
(218) YD Diagnostics URISCAN Pro+		>	1	0,3	7,5			
(220) 77 Elektronika HandUReader	CVP	>>>	1	0,3	8,0			
(221) Erba Lachema LAURA			1	0,3	6,5			
		>	6	2,1	7,0			
	CVP	>>>	11	3,8	8,0			
(222) Roche Urisys 1100		>	1	0,3	7,0			
	CVP	>>>	2	0,7	8,0			
(223) Iris iChem Velocity			1	0,3	5,5			
	CVP	>>>	13	4,5	8,0			
		>	9	3,1	8,5			
(225) Dirui H-series			2	0,7	6,5			
		>	2	0,7	7,0			
		>	3	1	7,5			
	CVP	>>>	2	0,7	8,0			
(228) Siemens Clinitek Advantus			1	0,3	6,5			
		>	3	1	7,0			
		>	1	0,3	7,5			
	CVP	>>>	4	1,4	8,0			
		>	1	0,3	8,5			
(229) 77 Elektronika LabUMat, LabUMat 2			>	7	2,4	7,5		
	CVP	>>>	20	6,8	8,0			
		>	4	1,4	8,5			
(230) Dirui FUS-series			>	1	0,3	7,0		
		>	33	11	7,5			
	CVP	>>>	19	6,5	8,0			
(998) Manual				1	0,3	6,0		
				2	0,7	6,5		
				>	6	2,1	7,0	
				>	6	2,1	7,5	
	CVP	>>>		9	3,1	8,0		
				>	1	0,3	8,5	

Summary statistics - qualitative results**EQA round: DGP1/24 - Urine Strip Tests Analysis**

Deadline: 16.2.2024

Setup: groups - S (measuring system)

Test Sample Group	AV	N _{tot}	N _{rel} [%]	Frequency of the results		N _{eva}	N _{suc}	S _{rel} [%]
				Result				
(999) another system	>	1	0,3	9,0				
		1	0,3	6,0				
	>	2	0,7	7,0				
	>	4	1,4	7,5				
	CVP	>>>	7	2,4	8,0			
(266) Urobilinogen		291				291	289	99
Sample A		291				291	289	99
(201) Arkray Aution JET / Eleven	CVP	>>>	5	1,7	Negative, < 17 µmol/L			
(202) Arkray Aution MAX / Sysmex UX-2000	CVP	>>>	10	3,4	Negative, < 17 µmol/L			
(204) Siemens Clinitek Novus	CVP	>>>	34	12	Negative, < 17 µmol/L			
(205) Roche cobas u 601 (cobas 6500)	CVP	>>>	9	3,1	Negative, < 17 µmol/L			
(208) Siemens Clinitek Status	CVP	>>>	4	1,4	Negative, < 17 µmol/L			
(209) 77 Elektronika LabUreader	CVP	>>>	10	3,4	Negative, < 17 µmol/L			
(215) Roche Urisys 1800 / cobas u411	CVP	>>>	12	4,1	Negative, < 17 µmol/L			
(216) Sysmex UC-1000, UC-3500	CVP	>>>	18	6,2	Negative, < 17 µmol/L			
(218) YD Diagnostics URISCAN Pro+	CVP	>>>	1	0,3	Negative, < 17 µmol/L			
(220) 77 Elektronika HandUReader	CVP	>>>	1	0,3	Negative, < 17 µmol/L			
(221) Erba Lachema LAURA	CVP	>>>	18	6,2	Negative, < 17 µmol/L			
(222) Roche Urisys 1100	CVP	>>>	3	1	Negative, < 17 µmol/L			
(223) Iris iChem Velocity	CVP	>>>	23	7,9	Negative, < 17 µmol/L			
(225) Dirui H-series	CVP	>>>	9	3,1	Negative, < 17 µmol/L			
(228) Siemens Clinitek Advantus	CVP	>>>	10	3,4	Negative, < 17 µmol/L			
(229) 77 Elektronika LabUMat, LabUMat 2	CVP	>>>	31	11	Negative, < 17 µmol/L			
(230) Dirui FUS-series	CVP	>>>	53	18	Negative, < 17 µmol/L			
(998) Manual	CVP	>>>	25	8,6	Negative, < 17 µmol/L			
			1	0,3	1+, 2+			
(999) another system	CVP	>>>	13	4,5	Negative, < 17 µmol/L			
			1	0,3	1+, 2+			
Sample B		291				291	290	100
(201) Arkray Aution JET / Eleven	CVP	>>>	5	1,7	Negative, < 17 µmol/L			
(202) Arkray Aution MAX / Sysmex UX-2000	CVP	>>>	10	3,4	Negative, < 17 µmol/L			
(204) Siemens Clinitek Novus	CVP	>>>	34	12	Negative, < 17 µmol/L			
(205) Roche cobas u 601 (cobas 6500)	CVP	>>>	9	3,1	Negative, < 17 µmol/L			
(208) Siemens Clinitek Status	CVP	>>>	4	1,4	Negative, < 17 µmol/L			
(209) 77 Elektronika LabUreader	CVP	>>>	10	3,4	Negative, < 17 µmol/L			
(215) Roche Urisys 1800 / cobas u411	CVP	>>>	12	4,1	Negative, < 17 µmol/L			
(216) Sysmex UC-1000, UC-3500	CVP	>>>	18	6,2	Negative, < 17 µmol/L			
(218) YD Diagnostics URISCAN Pro+	CVP	>>>	1	0,3	Negative, < 17 µmol/L			
(220) 77 Elektronika HandUReader	CVP	>>>	1	0,3	Negative, < 17 µmol/L			
(221) Erba Lachema LAURA	CVP	>>>	18	6,2	Negative, < 17 µmol/L			
(222) Roche Urisys 1100	CVP	>>>	3	1	Negative, < 17 µmol/L			
(223) Iris iChem Velocity	CVP	>>>	23	7,9	Negative, < 17 µmol/L			
(225) Dirui H-series	CVP	>>>	9	3,1	Negative, < 17 µmol/L			
(228) Siemens Clinitek Advantus	CVP	>>>	10	3,4	Negative, < 17 µmol/L			
(229) 77 Elektronika LabUMat, LabUMat 2	CVP	>>>	31	11	Negative, < 17 µmol/L			
(230) Dirui FUS-series	CVP	>>>	53	18	Negative, < 17 µmol/L			
(998) Manual	CVP	>>>	25	8,6	Negative, < 17 µmol/L			
			1	0,3	1+, 2+			
(999) another system	CVP	>>>	14	4,8	Negative, < 17 µmol/L			