Summary statistics - quantitative results

EQA round: INRP4/26 - INR measurement on POCT

Deadline: 23.10.2026

Setup: all data

RoM = robust average
SD = standard deviation
CV = coefficient of variation
N_{tot} = total number of the results
N_{out} = number of the results removed before calculation

AV = assigned value
CVP = consensus of all participants
U_{AV} = expanded uncertainty of the assigned value (k = 2)

D_{max} = acceptable difference
LL = lower limit
UL = upper limit
N_{eva} = number of the results assessed
N_{suc} = number of successful results
S_{rel} = relative success

Test / Sample / Group	[unit]	RoM	SD	CV [%]	N_{tot}	N_{out}	Comparability	AV	U_{AV}	D_{max}	LL	UL	N_{eva}	N_{suc}	S_{rel} [%]
(536) **Prothrombin test (CoaguChek)**	[-]				129								129	124	96
Sample A1													129	125	97
All results		4,04	0,2	5	129	0	CVP	4,04	0,044	asym.	3,2	4,9	129		
Sample B1													129	125	97
All results		1,76	0,063	3,6	129	0	CVP	1,76	0,014	asym.	1,4	2,2	129		