Summary statistics - qualitative results

EQA round: STA2/26 - Streptococcus A - identification

Deadline: 26.6.2026

Setup: groups - S (measuring system)

AV = assigned value
CVP = consensus of all participants
>>> ... expected result
> ... acceptable result
± ... result not assessed

N_{tot} = total number of the results
N_{rel} = relative number of the results

N_{eva} = number of the results assessed
N_{suc} = number of successful results
S_{rel} = relative success

Test / Sample / Group	AV	N_{tot}	N_{rel} [%]	Result	N_{eva}	N_{suc}	S_{rel} [%]
(287) **Streptococcus A - identification**		77			77	75	97
Sample A		77			77	75	97
(0) not specified	CVP >>>	5	6,5	Positive			
(130) Aidian QuikRead 101		1	1,3	Negative			
(134) Aidian QuikRead go		1	1,3	Negative			
	CVP >>>	66	86	Positive			
(139) SD Biosensor Standard F200	CVP >>>	3	3,9	Positive			
(999) another system	CVP >>>	1	1,3	Positive			
Sample B		77			77	76	99
(0) not specified	CVP >>>	5	6,5	Negative			
(130) Aidian QuikRead 101		1	1,3	Positive			
(134) Aidian QuikRead go	CVP >>>	67	87	Negative			
(139) SD Biosensor Standard F200	CVP >>>	3	3,9	Negative			
(999) another system	CVP >>>	1	1,3	Negative			