Summary statistics - quantitative results

EQA round: INRP2/26 - INR measurement on POCT

Deadline: 5.6.2026

Setup: all data

RoM = robust average
SD = standard deviation
CV = coefficient of variation
N_{tot} = total number of the results
N_{out} = number of the results removed before calculation

AV = assigned value
CVP = consensus of all participants
U_{AV} = expanded uncertainty of the assigned value (k = 2)

D_{max} = acceptable difference
LL = lower limit
UL = upper limit
N_{eva} = number of the results assessed
N_{suc} = number of successful results
S_{rel} = relative success

Test / Sample / Group	[unit]	RoM	SD	CV [%]	N_{tot}	N_{out}	Comparability	AV	U_{AV}	D_{max}	LL	UL	N_{eva}	N_{suc}	S_{rel} [%]
(536) **Prothrombin test (CoaguChek)**	[-]				71								71	70	99
Sample A1													71	70	99
All results		1,76	0,062	3,5	71	0	CVP	1,76	0,018	20%	1,4	2,12	71		
Sample B1													71	70	99
All results		2,27	0,16	7	71	0	CVP	2,27	0,047	20%	1,81	2,73	71		